

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3141</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047749.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0923WBSD03	Dichlorodifluoromethane	20	19.6	ug/L	98	19		32	152	20
	Chloromethane	20	18.5	ug/L	93	16		50	139	20
	Vinyl chloride	20	18.4	ug/L	92	20		58	137	20
	Bromomethane	20	19.7	ug/L	99	17		53	141	20
	Chloroethane	20	17.6	ug/L	88	18		60	138	20
	Trichlorofluoromethane	20	17.6	ug/L	88	17		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	17.3	ug/L	86	20		70	136	20
	1,1-Dichloroethene	20	18.0	ug/L	90	16		71	131	20
	Acetone	100	75.9	ug/L	76	15		39	160	20
	Carbon disulfide	20	18.0	ug/L	90	14		64	133	20
	Methyl tert-butyl Ether	20	17.5	ug/L	88	16		71	124	20
	Methyl Acetate	20	19.8	ug/L	99	12		56	136	20
	Methylene Chloride	20	17.7	ug/L	89	16		74	124	20
	trans-1,2-Dichloroethene	20	17.8	ug/L	89	18		75	124	20
	1,1-Dichloroethane	20	17.6	ug/L	88	18		77	125	20
	Cyclohexane	20	17.4	ug/L	87	18		71	130	20
	2-Butanone	100	90.8	ug/L	91	19		56	143	20
	Carbon Tetrachloride	20	16.5	ug/L	83	18		72	136	20
	cis-1,2-Dichloroethene	20	17.9	ug/L	90	14		78	123	20
	Bromochloromethane	20	21.7	ug/L	109	13		78	123	20
	Chloroform	20	17.9	ug/L	90	16		79	124	20
	1,1,1-Trichloroethane	20	18.0	ug/L	90	16		74	131	20
	Methylcyclohexane	20	16.6	ug/L	83	20		72	132	20
	Benzene	20	17.7	ug/L	89	16		79	120	20
	1,2-Dichloroethane	20	16.8	ug/L	84	19		73	128	20
	Trichloroethene	20	17.9	ug/L	90	16		79	123	20
	1,2-Dichloropropane	20	17.9	ug/L	90	15		78	122	20
	Bromodichloromethane	20	17.5	ug/L	88	17		79	125	20
	4-Methyl-2-Pentanone	100	96.4	ug/L	96	14		67	130	20
	Toluene	20	17.7	ug/L	89	17		80	121	20
	t-1,3-Dichloropropene	20	16.0	ug/L	80	16		73	127	20
	cis-1,3-Dichloropropene	20	16.0	ug/L	80	18		75	124	20
	1,1,2-Trichloroethane	20	18.2	ug/L	91	18		80	119	20
	2-Hexanone	100	94.0	ug/L	94	16		57	139	20
	Dibromochloromethane	20	18.1	ug/L	91	15		74	126	20
	1,2-Dibromoethane	20	19.1	ug/L	96	14		77	121	20
	Tetrachloroethene	20	18.5	ug/L	93	14		74	129	20
	Chlorobenzene	20	17.7	ug/L	89	16		82	118	20
	Ethyl Benzene	20	17.2	ug/L	86	19		79	121	20
	m/p-Xylenes	40	35.3	ug/L	88	19		80	121	20
	o-Xylene	20	17.5	ug/L	88	17		78	122	20
	Styrene	20	17.5	ug/L	88	17		78	123	20
	Bromoform	20	17.5	ug/L	88	17		66	130	20
	Isopropylbenzene	20	16.9	ug/L	85	19		72	131	20
	1,1,2,2-Tetrachloroethane	20	18.5	ug/L	93	15		71	121	20
	1,3-Dichlorobenzene	20	17.4	ug/L	87	17		80	119	20
	1,4-Dichlorobenzene	20	17.1	ug/L	86	18		79	118	20
	1,2-Dichlorobenzene	20	17.5	ug/L	88	17		80	119	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3141 Analytical Method: SW8260-Low
Client: Tetra Tech NUS, Inc. Datafile : VX047749.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0923WBSD03	1,2-Dibromo-3-Chloropropane	20	17.6	ug/L	88	15		62	128	20
	1,2,4-Trichlorobenzene	20	16.4	ug/L	82	19		69	130	20
	1,2,3-Trichlorobenzene	20	17.1	ug/L	86	19		69	129	20