

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3141
Lab File ID: VX047746.D Date Analyzed: 09/24/2025
Instrument ID: MSVOA_X Time Analyzed: 02:08
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	166745	5.53	290663	6.74	252770	10.04
UPPER LIMIT	333490	6.032	581326	7.239	505540	10.537
LOWER LIMIT	83372.5	5.032	145332	6.239	126385	9.537
EPA SAMPLE NO.						
BP-VPB-184-EB-20250918	157285	5.54	293648	6.75	269079	10.04
BP-VPB-184-GW-100-102	165497	5.54	300349	6.75	276860	10.04
VX0923WBL02	197095	5.54	377977	6.75	356070	10.04
VX0923WBS03	156932	5.54	284011	6.75	253957	10.04
VX0923WBSD03	154222	5.54	281988	6.75	254297	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.