

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q3141
 Lab File ID: VX047719.D Date Analyzed: 09/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	115061	12.00€				
UPPER LIMIT	230122	12.50€				
LOWER LIMIT	57530.5	11.50€				
EPA SAMPLE NO.						
BP-VPB-184-TB-20250917	128339	12.01				
BP-VPB-184-GW-60-62	155521	12.01				
BP-VPB-184-GW-150-152	153459	12.01				
BP-VPB-184-GW-200-202	146402	12.01				
BP-VPB-184-GW-220-222	157643	12.01				
VX0923WBL01	132951	12.01				
VX0923WBS02	141661	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.