

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06  
 Lab Code: ACE SDG NO.: Q3141  
 Lab File ID: VX047746.D Date Analyzed: 09/24/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 02:08  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                        | IS4<br>AREA # | RT #   |  |  |  |  |
|------------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD            | 121948        | 12.00€ |  |  |  |  |
| UPPER LIMIT            | 243896        | 12.50€ |  |  |  |  |
| LOWER LIMIT            | 60974         | 11.50€ |  |  |  |  |
| EPA SAMPLE NO.         |               |        |  |  |  |  |
| BP-VPB-184-EB-20250918 | 126784        | 12.01  |  |  |  |  |
| BP-VPB-184-GW-100-102  | 136496        | 12.01  |  |  |  |  |
| VX0923WBL02            | 171478        | 12.01  |  |  |  |  |
| VX0923WBS03            | 123765        | 12.01  |  |  |  |  |
| VX0923WBSD03           | 124629        | 12.01  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.