

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3141

Lab File ID: VX047745.D

BFB Injection Date: 09/24/2025

Instrument ID: MSVOA_X

BFB Injection Time: 01:30

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.9
75	30.0 - 60.0% of mass 95	51.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	68
175	5.0 - 9.0% of mass 174	5 (7.4) 1
176	95.0 - 101.0% of mass 174	65.9 (97) 1
177	5.0 - 9.0% of mass 176	4.3 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047746.D	09/24/2025	02:08
VX0923WBL02	VX0923WBL02	VX047747.D	09/24/2025	03:11
VX0923WBS03	VX0923WBS03	VX047748.D	09/24/2025	03:53
VX0923WBSD03	VX0923WBSD03	VX047749.D	09/24/2025	04:14
BP-VPB-184-GW-100-102	Q3141-04	VX047760.D	09/24/2025	08:22
BP-VPB-184-EB-20250918	Q3141-02	VX047761.D	09/24/2025	08:43
VSTDCCC050EC	VSTDCCC050	VX047767.D	09/24/2025	10:50