

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q3142

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB137243

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137243BS							
TDS	mg/L	100	94.0		94	1	90-110	09/19/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q3142

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB137250

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137250BS							
TSS	mg/L	550	532		97	1	90-110	09/19/2025