

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3142

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3960	4000	99	90 - 110	P	09/24/2025	11:34	LB137295

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3142

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	115	100	115	80 - 120	P	09/24/2025	11:44	LB137295

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3142

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5200	5000	104	90 - 110	P	09/24/2025	12:21	LB137295
CCV02	Iron	5020	5000	100	90 - 110	P	09/24/2025	13:40	LB137295
CCV03	Iron	4740	5000	95	90 - 110	P	09/24/2025	14:42	LB137295
CCV04	Iron	4960	5000	99	90 - 110	P	09/24/2025	15:41	LB137295
CCV05	Iron	4990	5000	100	90 - 110	P	09/24/2025	16:31	LB137295
CCV06	Iron	4900	5000	98	90 - 110	P	09/24/2025	17:33	LB137295
CCV07	Iron	4850	5000	97	90 - 110	P	09/24/2025	18:37	LB137295
CCV08	Iron	4890	5000	98	90 - 110	P	09/24/2025	19:03	LB137295