

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc.

SDG No.: Q3142

Contract: TETR06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	103000	101000	102	85600	116500	09/24/2025	12:04	LB137295
ICSAB01	Iron	103000	99300	104	84400	114500	09/24/2025	12:08	LB137295