

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3143

Client ID : SSTD020383

Date Analyzed: 10/03/2025

Lab File ID: BN037998.D

Time Analyzed: 15:27

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	273492	7.805	1249710	10.61	826560	14.46
UPPER LIMIT	546984	8.305	2499420	11.105	1653120	14.963
LOWER LIMIT	136746	7.305	624855	10.105	413280	13.963
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	268772	7.81	1129910	10.61	705743	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.