

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	09/12/25
Project:	Weekly Storage Blanks		Date Received:	09/19/25
Client Sample ID:	STORAGE-BLANK-SAMPLE-MANAGEMENT		SDG No.:	Q3143
Lab Sample ID:	Q3143-04		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC- Chemtech Full
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047861.D	1	09/26/25 16:10	VX092625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
141-78-6	Ethyl Acetate	0.31	U	0.31	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-65-0	Tert butyl alcohol	5.50	U	5.50	25.0	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
107-02-8	Acrolein	7.10	U	7.10	25.0	ug/L
107-13-1	Acrylonitrile	0.83	U	0.83	5.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
108-05-4	Vinyl Acetate	0.66	U	0.66	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
594-20-7	2,2-Dichloropropane	0.21	U	0.21	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
563-58-6	1,1-Dichloropropene	0.15	U	0.15	1.00	ug/L

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	09/12/25	
Project:	Weekly Storage Blanks		Date Received:	09/19/25	
Client Sample ID:	STORAGE-BLANK-SAMPLE-MANAGEMENT		SDG No.:	Q3143	
Lab Sample ID:	Q3143-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC- Chemtech Full	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047861.D	1	09/26/25 16:10	VX092625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
74-95-3	Dibromomethane	0.25	U	0.25	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	UQ	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
142-28-9	1,3-Dichloropropane	0.19	U	0.19	1.00	ug/L
110-75-8	2-Chloroethyl Vinyl ether	0.30	U	0.30	5.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	UQ	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
630-20-6	1,1,1,2-Tetrachloroethane	0.19	U	0.19	1.00	ug/L
67-72-1	Hexachloroethane	0.32	U	0.32	0	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
96-18-4	1,2,3-Trichloropropane	0.35	U	0.35	1.00	ug/L
108-86-1	Bromobenzene	0.24	U	0.24	1.00	ug/L
103-65-1	n-propylbenzene	0.13	U	0.13	1.00	ug/L
95-49-8	2-Chlorotoluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	09/12/25	
Project:	Weekly Storage Blanks		Date Received:	09/19/25	
Client Sample ID:	STORAGE-BLANK-SAMPLE-MANAGEMENT		SDG No.:	Q3143	
Lab Sample ID:	Q3143-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC- Chemtech Full	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047861.D	1	09/26/25 16:10	VX092625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
108-67-8	1,3,5-Trimethylbenzene	0.15	U	0.15	1.00	ug/L
106-43-4	4-Chlorotoluene	0.13	U	0.13	1.00	ug/L
98-06-6	tert-Butylbenzene	0.14	U	0.14	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.14	U	0.14	1.00	ug/L
135-98-8	sec-Butylbenzene	0.13	U	0.13	1.00	ug/L
99-87-6	p-Isopropyltoluene	0.13	U	0.13	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
104-51-8	n-Butylbenzene	0.15	U	0.15	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-68-3	Hexachlorobutadiene	0.30	U	0.30	1.00	ug/L
91-20-3	Naphthalene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.8		74 - 125	102%	SPK: 50
1868-53-7	Dibromofluoromethane	48.7		75 - 124	97%	SPK: 50
2037-26-5	Toluene-d8	48.9		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.0		77 - 121	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	118000	5.544			
540-36-3	1,4-Difluorobenzene	244000	6.751			
3114-55-4	Chlorobenzene-d5	252000	10.037			
3855-82-1	1,4-Dichlorobenzene-d4	122000	12.006			

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	09/12/25
Project:	Weekly Storage Blanks		Date Received:	09/19/25
Client Sample ID:	STORAGE-BLANK-SAMPLE-MANAGEMENT		SDG No.:	Q3143
Lab Sample ID:	Q3143-04		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC- Chemtech Full
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047861.D	1	09/26/25 16:10	VX092625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products