

Surrogate Summary

SDG No.: **Q3143**

Client: **Alliance Technical Group, LLC - Newark**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3143-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	51.7	103		74	125
		Dibromofluoromethane	50	48.8	98		75	124
		Toluene-d8	50	49.2	98		86	113
		4-Bromofluorobenzene	50	52.7	105		77	121
Q3143-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	52.5	105		74	125
		Dibromofluoromethane	50	49.7	99		75	124
		Toluene-d8	50	48.9	98		86	113
		4-Bromofluorobenzene	50	53.8	108		77	121
Q3143-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	51.4	103		74	125
		Dibromofluoromethane	50	48.2	96		75	124
		Toluene-d8	50	48.4	97		86	113
		4-Bromofluorobenzene	50	53.0	106		77	121
Q3143-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	50.8	102		74	125
		Dibromofluoromethane	50	48.7	97		75	124
		Toluene-d8	50	48.9	98		86	113
		4-Bromofluorobenzene	50	53.0	106		77	121
VX0926WBL01	VX0926WBL01	1,2-Dichloroethane-d4	50	50.4	101		74	125
		Dibromofluoromethane	50	48.4	97		75	124
		Toluene-d8	50	49.3	99		86	113
		4-Bromofluorobenzene	50	53.2	106		77	121
VX0926WBS01	VX0926WBS01	1,2-Dichloroethane-d4	50	45.7	91		74	125
		Dibromofluoromethane	50	50.9	102		75	124
		Toluene-d8	50	51.6	103		86	113
		4-Bromofluorobenzene	50	51.4	103		77	121
VX0926WBSD0	VX0926WBSD01	1,2-Dichloroethane-d4	50	48.4	97		74	125
		Dibromofluoromethane	50	51.8	104		75	124
		Toluene-d8	50	50.8	102		86	113
		4-Bromofluorobenzene	50	51.7	103		77	121