

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG NO.: Q3143
 Lab File ID: VX047848.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:54
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	121207	12.00€				
UPPER LIMIT	242414	12.50€				
LOWER LIMIT	60603.5	11.50€				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	120344	12.01				
STORAGE-BLANK-WATER-REF-5	134741	12.01				
STORAGE-BLANK-WATER-REF-4	124194	12.01				
STORAGE-BLANK-SAMPLE-MANAGEMENT	122127	12.01				
VX0926WBL01	183902	12.01				
VX0926WBS01	120010	12.01				
VX0926WBSD01	112641	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.