

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q3144

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3850	4000	96	95 - 105	P	09/22/2025	12:31	LB137271

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q3144

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	12.3	12.0	102	80 - 120	P	09/22/2025	12:48	LB137271

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q3144

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4930	5000	99	90 - 110	P	09/22/2025	13:36	LB137271
CCV02	Lead	4880	5000	98	90 - 110	P	09/22/2025	14:48	LB137271
CCV03	Lead	4790	5000	96	90 - 110	P	09/22/2025	15:46	LB137271
CCV04	Lead	4790	5000	96	90 - 110	P	09/22/2025	16:27	LB137271
CCV05	Lead	5020	5000	100	90 - 110	P	09/22/2025	16:43	LB137271
CCV06	Lead	4890	5000	98	90 - 110	P	09/22/2025	17:49	LB137271
CCV07	Lead	5070	5000	101	90 - 110	P	09/22/2025	18:49	LB137271