

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q3144

Contract: SAFE04

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	-6.20			-12	12	09/22/2025	13:10	LB137271
ICSAB01	Lead	45.3	49.0	92	37	61	09/22/2025	13:14	LB137271
ICSA	Lead	22.4			-12	12	09/22/2025	13:22	LB137271
ICSAB	Lead	57.9	49.0	118	37	61	09/22/2025	13:26	LB137271