

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q3144

Contract: SAFE04

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169779BS Lead	ug/L	500	477		95	85 - 115	P