

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3148

Project: Transfer Station-SPDES

Run No.: LB137250

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137250BS							
TSS	mg/L	550	532		97	1	90-110	09/19/2025