

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3148

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1020	1000	102	95 - 105	P	09/22/2025	12:31	LB137271
	Iron	3960	4000	99	95 - 105	P	09/22/2025	12:31	LB137271
	Lead	3850	4000	96	95 - 105	P	09/22/2025	12:31	LB137271

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3148

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.3	20.0	111	80 - 120	P	09/22/2025	12:48	LB137271
	Iron	83.7	100	84	80 - 120	P	09/22/2025	12:48	LB137271
	Lead	12.3	12.0	102	80 - 120	P	09/22/2025	12:48	LB137271

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3148

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	09/22/2025	13:36	LB137271
	Iron	5000	5000	100	90 - 110	P	09/22/2025	13:36	LB137271
	Lead	4930	5000	99	90 - 110	P	09/22/2025	13:36	LB137271
CCV02	Copper	1230	1250	99	90 - 110	P	09/22/2025	14:48	LB137271
	Iron	4990	5000	100	90 - 110	P	09/22/2025	14:48	LB137271
	Lead	4880	5000	98	90 - 110	P	09/22/2025	14:48	LB137271
CCV03	Copper	1210	1250	97	90 - 110	P	09/22/2025	15:46	LB137271
	Iron	4830	5000	96	90 - 110	P	09/22/2025	15:46	LB137271
	Lead	4790	5000	96	90 - 110	P	09/22/2025	15:46	LB137271
CCV04	Copper	1220	1250	97	90 - 110	P	09/22/2025	16:27	LB137271
	Iron	4930	5000	99	90 - 110	P	09/22/2025	16:27	LB137271
	Lead	4790	5000	96	90 - 110	P	09/22/2025	16:27	LB137271
CCV05	Copper	1270	1250	102	90 - 110	P	09/22/2025	16:43	LB137271
	Iron	5050	5000	101	90 - 110	P	09/22/2025	16:43	LB137271
	Lead	5020	5000	100	90 - 110	P	09/22/2025	16:43	LB137271
CCV06	Copper	1240	1250	99	90 - 110	P	09/22/2025	17:49	LB137271
	Iron	4960	5000	99	90 - 110	P	09/22/2025	17:49	LB137271
	Lead	4890	5000	98	90 - 110	P	09/22/2025	17:49	LB137271
CCV07	Copper	1280	1250	102	90 - 110	P	09/22/2025	18:49	LB137271
	Iron	5150	5000	103	90 - 110	P	09/22/2025	18:49	LB137271
	Lead	5070	5000	101	90 - 110	P	09/22/2025	18:49	LB137271