

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q3148

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	13.2	2.0	660	-18	22	09/22/2025	13:10	LB137271
	Iron	109000	101000	108	85600	116500	09/22/2025	13:10	LB137271
	Lead	-6.20			-12	12	09/22/2025	13:10	LB137271
ICSAB01	Copper	521	511	102	434	588	09/22/2025	13:14	LB137271
	Iron	104000	99300	105	84400	114500	09/22/2025	13:14	LB137271
	Lead	45.3	49.0	92	37	61	09/22/2025	13:14	LB137271
ICSA	Copper	5.60	2.0	280	-18	22	09/22/2025	13:22	LB137271
	Iron	91800	101000	91	85600	116500	09/22/2025	13:22	LB137271
	Lead	22.4			-12	12	09/22/2025	13:22	LB137271
ICSAB	Copper	532	511	104	434	588	09/22/2025	13:26	LB137271
	Iron	99300	99300	100	84400	114500	09/22/2025	13:26	LB137271
	Lead	57.9	49.0	118	37	61	09/22/2025	13:26	LB137271