

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV74	Mercury	4.15	4.0	104	90 - 110	CV	10/16/2025	09:36	LB137552

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV37	Mercury	5.13	5.0	102	90 - 110	CV	10/16/2025	09:41	LB137552
CCV38	Mercury	5.10	5.0	102	90 - 110	CV	10/16/2025	10:13	LB137552
CCV39	Mercury	5.14	5.0	103	90 - 110	CV	10/16/2025	10:28	LB137552

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3950	4000	99	90 - 110	P	10/15/2025	14:07	LB137561
	Barium	7910	8000	99	90 - 110	P	10/15/2025	14:07	LB137561
	Cadmium	1930	2000	96	90 - 110	P	10/15/2025	14:07	LB137561
	Chromium	795	800	99	90 - 110	P	10/15/2025	14:07	LB137561
	Lead	3800	4000	95	90 - 110	P	10/15/2025	14:07	LB137561
	Selenium	4020	4000	101	90 - 110	P	10/15/2025	14:07	LB137561
	Silver	953	1000	95	90 - 110	P	10/15/2025	14:07	LB137561

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	21.8	20.0	109	80 - 120	P	10/15/2025	14:19	LB137561
	Barium	92.9	100	93	80 - 120	P	10/15/2025	14:19	LB137561
	Cadmium	5.77	6.0	96	80 - 120	P	10/15/2025	14:19	LB137561
	Chromium	10.4	10.0	104	80 - 120	P	10/15/2025	14:19	LB137561
	Lead	12.2	12.0	102	80 - 120	P	10/15/2025	14:19	LB137561
	Selenium	19.1	20.0	95	80 - 120	P	10/15/2025	14:19	LB137561
	Silver	10.7	10.0	107	80 - 120	P	10/15/2025	14:19	LB137561

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5000	5000	100	90 - 110	P	10/15/2025	15:09	LB137561
	Barium	9940	10000	99	90 - 110	P	10/15/2025	15:09	LB137561
	Cadmium	2480	2500	99	90 - 110	P	10/15/2025	15:09	LB137561
	Chromium	1010	1000	101	90 - 110	P	10/15/2025	15:09	LB137561
	Lead	4950	5000	99	90 - 110	P	10/15/2025	15:09	LB137561
	Selenium	5050	5000	101	90 - 110	P	10/15/2025	15:09	LB137561
	Silver	1250	1250	100	90 - 110	P	10/15/2025	15:09	LB137561
CCV02	Arsenic	4810	5000	96	90 - 110	P	10/15/2025	16:28	LB137561
	Barium	9240	10000	92	90 - 110	P	10/15/2025	16:28	LB137561
	Cadmium	2340	2500	93	90 - 110	P	10/15/2025	16:28	LB137561
	Chromium	948	1000	95	90 - 110	P	10/15/2025	16:28	LB137561
	Lead	4690	5000	94	90 - 110	P	10/15/2025	16:28	LB137561
	Selenium	4910	5000	98	90 - 110	P	10/15/2025	16:28	LB137561
	Silver	1200	1250	96	90 - 110	P	10/15/2025	16:28	LB137561
CCV03	Arsenic	4930	5000	98	90 - 110	P	10/15/2025	17:21	LB137561
	Barium	9320	10000	93	90 - 110	P	10/15/2025	17:21	LB137561
	Cadmium	2410	2500	96	90 - 110	P	10/15/2025	17:21	LB137561
	Chromium	988	1000	99	90 - 110	P	10/15/2025	17:21	LB137561
	Lead	4840	5000	97	90 - 110	P	10/15/2025	17:21	LB137561
	Selenium	5060	5000	101	90 - 110	P	10/15/2025	17:21	LB137561
	Silver	1230	1250	98	90 - 110	P	10/15/2025	17:21	LB137561
CCV04	Arsenic	4750	5000	95	90 - 110	P	10/15/2025	18:10	LB137561
	Barium	9160	10000	92	90 - 110	P	10/15/2025	18:10	LB137561
	Cadmium	2360	2500	94	90 - 110	P	10/15/2025	18:10	LB137561
	Chromium	976	1000	98	90 - 110	P	10/15/2025	18:10	LB137561
	Lead	4710	5000	94	90 - 110	P	10/15/2025	18:10	LB137561
	Selenium	4800	5000	96	90 - 110	P	10/15/2025	18:10	LB137561
	Silver	1200	1250	96	90 - 110	P	10/15/2025	18:10	LB137561
CCV05	Arsenic	4940	5000	99	90 - 110	P	10/15/2025	18:58	LB137561
	Barium	9690	10000	97	90 - 110	P	10/15/2025	18:58	LB137561
	Cadmium	2480	2500	99	90 - 110	P	10/15/2025	18:58	LB137561
	Chromium	1020	1000	102	90 - 110	P	10/15/2025	18:58	LB137561
	Lead	4940	5000	99	90 - 110	P	10/15/2025	18:58	LB137561
	Selenium	4960	5000	99	90 - 110	P	10/15/2025	18:58	LB137561

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1260	1250	101	90 - 110	P	10/15/2025	18:58	LB137561
CCV06	Arsenic	4940	5000	99	90 - 110	P	10/15/2025	19:44	LB137561
	Barium	10100	10000	101	90 - 110	P	10/15/2025	19:44	LB137561
	Cadmium	2460	2500	98	90 - 110	P	10/15/2025	19:44	LB137561
	Chromium	1020	1000	102	90 - 110	P	10/15/2025	19:44	LB137561
	Lead	4900	5000	98	90 - 110	P	10/15/2025	19:44	LB137561
	Selenium	4970	5000	100	90 - 110	P	10/15/2025	19:44	LB137561
	Silver	1240	1250	100	90 - 110	P	10/15/2025	19:44	LB137561

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3920	4000	98	95 - 105	P	10/16/2025	11:34	LB137573
	Barium	7880	8000	98	95 - 105	P	10/16/2025	11:34	LB137573
	Cadmium	1930	2000	97	95 - 105	P	10/16/2025	11:34	LB137573
	Chromium	797	800	100	95 - 105	P	10/16/2025	11:34	LB137573
	Lead	3800	4000	95	95 - 105	P	10/16/2025	11:34	LB137573
	Selenium	3910	4000	98	95 - 105	P	10/16/2025	11:34	LB137573
	Silver	965	1000	96	95 - 105	P	10/16/2025	11:34	LB137573

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	20.2	20.0	101	80 - 120	P	10/16/2025	11:58	LB137573
	Barium	92.6	100	93	80 - 120	P	10/16/2025	11:58	LB137573
	Cadmium	5.80	6.0	97	80 - 120	P	10/16/2025	11:58	LB137573
	Chromium	10.3	10.0	103	80 - 120	P	10/16/2025	11:58	LB137573
	Lead	10.6	12.0	89	80 - 120	P	10/16/2025	11:58	LB137573
	Selenium	21.3	20.0	107	80 - 120	P	10/16/2025	11:58	LB137573
	Silver	10.4	10.0	104	80 - 120	P	10/16/2025	11:58	LB137573

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5080	5000	102	90 - 110	P	10/16/2025	12:38	LB137573
	Barium	9890	10000	99	90 - 110	P	10/16/2025	12:38	LB137573
	Cadmium	2510	2500	100	90 - 110	P	10/16/2025	12:38	LB137573
	Chromium	1010	1000	101	90 - 110	P	10/16/2025	12:38	LB137573
	Lead	4970	5000	100	90 - 110	P	10/16/2025	12:38	LB137573
	Selenium	5110	5000	102	90 - 110	P	10/16/2025	12:38	LB137573
	Silver	1250	1250	100	90 - 110	P	10/16/2025	12:38	LB137573
CCV02	Arsenic	4920	5000	98	90 - 110	P	10/16/2025	13:45	LB137573
	Barium	9950	10000	100	90 - 110	P	10/16/2025	13:45	LB137573
	Cadmium	2390	2500	96	90 - 110	P	10/16/2025	13:45	LB137573
	Chromium	958	1000	96	90 - 110	P	10/16/2025	13:45	LB137573
	Lead	4800	5000	96	90 - 110	P	10/16/2025	13:45	LB137573
	Selenium	4940	5000	99	90 - 110	P	10/16/2025	13:45	LB137573
	Silver	1220	1250	98	90 - 110	P	10/16/2025	13:45	LB137573
CCV03	Arsenic	4840	5000	97	90 - 110	P	10/16/2025	14:45	LB137573
	Barium	9620	10000	96	90 - 110	P	10/16/2025	14:45	LB137573
	Cadmium	2360	2500	95	90 - 110	P	10/16/2025	14:45	LB137573
	Chromium	959	1000	96	90 - 110	P	10/16/2025	14:45	LB137573
	Lead	4740	5000	95	90 - 110	P	10/16/2025	14:45	LB137573
	Selenium	4870	5000	97	90 - 110	P	10/16/2025	14:45	LB137573
	Silver	1210	1250	97	90 - 110	P	10/16/2025	14:45	LB137573
CCV04	Arsenic	4880	5000	98	90 - 110	P	10/16/2025	15:38	LB137573
	Barium	9380	10000	94	90 - 110	P	10/16/2025	15:38	LB137573
	Cadmium	2370	2500	95	90 - 110	P	10/16/2025	15:38	LB137573
	Chromium	975	1000	98	90 - 110	P	10/16/2025	15:38	LB137573
	Lead	4760	5000	95	90 - 110	P	10/16/2025	15:38	LB137573
	Selenium	4970	5000	99	90 - 110	P	10/16/2025	15:38	LB137573
	Silver	1220	1250	98	90 - 110	P	10/16/2025	15:38	LB137573
CCV05	Arsenic	5080	5000	102	90 - 110	P	10/16/2025	16:31	LB137573
	Barium	9930	10000	99	90 - 110	P	10/16/2025	16:31	LB137573
	Cadmium	2400	2500	96	90 - 110	P	10/16/2025	16:31	LB137573
	Chromium	964	1000	96	90 - 110	P	10/16/2025	16:31	LB137573
	Lead	4880	5000	98	90 - 110	P	10/16/2025	16:31	LB137573
	Selenium	5210	5000	104	90 - 110	P	10/16/2025	16:31	LB137573

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1260	1250	101	90 - 110	P	10/16/2025	16:31	LB137573
CCV06	Arsenic	4810	5000	96	90 - 110	P	10/16/2025	17:49	LB137573
	Barium	9330	10000	93	90 - 110	P	10/16/2025	17:49	LB137573
	Cadmium	2300	2500	92	90 - 110	P	10/16/2025	17:49	LB137573
	Chromium	932	1000	93	90 - 110	P	10/16/2025	17:49	LB137573
	Lead	4650	5000	93	90 - 110	P	10/16/2025	17:49	LB137573
	Selenium	4920	5000	98	90 - 110	P	10/16/2025	17:49	LB137573
	Silver	1200	1250	96	90 - 110	P	10/16/2025	17:49	LB137573
CCV07	Arsenic	4900	5000	98	90 - 110	P	10/16/2025	18:42	LB137573
	Barium	9660	10000	97	90 - 110	P	10/16/2025	18:42	LB137573
	Cadmium	2350	2500	94	90 - 110	P	10/16/2025	18:42	LB137573
	Chromium	956	1000	96	90 - 110	P	10/16/2025	18:42	LB137573
	Lead	4760	5000	95	90 - 110	P	10/16/2025	18:42	LB137573
	Selenium	4950	5000	99	90 - 110	P	10/16/2025	18:42	LB137573
	Silver	1220	1250	98	90 - 110	P	10/16/2025	18:42	LB137573
CCV08	Arsenic	4970	5000	99	90 - 110	P	10/16/2025	19:33	LB137573
	Barium	9350	10000	94	90 - 110	P	10/16/2025	19:33	LB137573
	Cadmium	2370	2500	95	90 - 110	P	10/16/2025	19:33	LB137573
	Chromium	971	1000	97	90 - 110	P	10/16/2025	19:33	LB137573
	Lead	4800	5000	96	90 - 110	P	10/16/2025	19:33	LB137573
	Selenium	5120	5000	102	90 - 110	P	10/16/2025	19:33	LB137573
	Silver	1230	1250	99	90 - 110	P	10/16/2025	19:33	LB137573
CCV09	Arsenic	4740	5000	95	90 - 110	P	10/16/2025	20:07	LB137573
	Barium	8960	10000	90	90 - 110	P	10/16/2025	20:07	LB137573
	Cadmium	2280	2500	91	90 - 110	P	10/16/2025	20:07	LB137573
	Chromium	939	1000	94	90 - 110	P	10/16/2025	20:07	LB137573
	Lead	4600	5000	92	90 - 110	P	10/16/2025	20:07	LB137573
	Selenium	4800	5000	96	90 - 110	P	10/16/2025	20:07	LB137573
	Silver	1190	1250	95	90 - 110	P	10/16/2025	20:07	LB137573