



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	25.7	20.0	129	65 - 135	P	10/15/2025	14:28	LB137561
	Barium	103	100	103	65 - 135	P	10/15/2025	14:28	LB137561
	Cadmium	6.21	6.0	103	65 - 135	P	10/15/2025	14:28	LB137561
	Chromium	11.2	10.0	112	65 - 135	P	10/15/2025	14:28	LB137561
	Lead	13.1	12.0	109	65 - 135	P	10/15/2025	14:28	LB137561
	Selenium	22.1	20.0	110	65 - 135	P	10/15/2025	14:28	LB137561
	Silver	11.5	10.0	115	65 - 135	P	10/15/2025	14:28	LB137561
CRA	Mercury	0.19	0.2	94	70 - 130	CV	10/16/2025	09:46	LB137552
CRI01	Arsenic	20.4	20.0	102	65 - 135	P	10/16/2025	12:09	LB137573
	Barium	98.6	100	99	65 - 135	P	10/16/2025	12:09	LB137573
	Cadmium	5.97	6.0	100	65 - 135	P	10/16/2025	12:09	LB137573
	Chromium	10.9	10.0	109	65 - 135	P	10/16/2025	12:09	LB137573
	Lead	10.5	12.0	88	65 - 135	P	10/16/2025	12:09	LB137573
	Selenium	21.5	20.0	107	65 - 135	P	10/16/2025	12:09	LB137573
	Silver	10.8	10.0	108	65 - 135	P	10/16/2025	12:09	LB137573