



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB74	Mercury	0.076	+/-0.2	U	0.20	CV	10/16/2025	09:38	LB137552

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB37	Mercury	0.076	+/-0.2	U	0.20	CV	10/16/2025	09:43	LB137552
CCB38	Mercury	0.076	+/-0.2	U	0.20	CV	10/16/2025	10:16	LB137552
CCB39	Mercury	0.076	+/-0.2	U	0.20	CV	10/16/2025	10:31	LB137552

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	14:24	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	14:24	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	14:24	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	14:24	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	14:24	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	14:24	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	14:24	LB137561

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	15:13	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	15:13	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	15:13	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	15:13	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	15:13	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	15:13	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	15:13	LB137561
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	16:32	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	16:32	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	16:32	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	16:32	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	16:32	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	16:32	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	16:32	LB137561
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	17:25	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	17:25	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	17:25	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	17:25	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	17:25	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	17:25	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	17:25	LB137561
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	18:14	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	18:14	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	18:14	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	18:14	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	18:14	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	18:14	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	18:14	LB137561
CCB05	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	19:02	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	19:02	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	19:02	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	19:02	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	19:02	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	19:02	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	19:02	LB137561
CCB06	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	19:48	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	19:48	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	19:48	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	19:48	LB137561

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	10/15/2025	19:48	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	19:48	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	19:48	LB137561

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	12:02	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	12:02	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	12:02	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	12:02	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	12:02	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	12:02	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	12:02	LB137573

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	12:42	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	12:42	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	12:42	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	12:42	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	12:42	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	12:42	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	12:42	LB137573
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	13:49	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	13:49	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	13:49	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	13:49	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	13:49	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	13:49	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	13:49	LB137573
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	14:49	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	14:49	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	14:49	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	14:49	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	14:49	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	14:49	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	14:49	LB137573
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	15:43	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	15:43	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	15:43	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	15:43	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	15:43	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	15:43	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	15:43	LB137573
CCB05	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	16:35	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	16:35	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	16:35	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	16:35	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	16:35	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	16:35	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	16:35	LB137573
CCB06	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	17:53	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	17:53	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	17:53	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	17:53	LB137573

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	10/16/2025	17:53	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	17:53	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	17:53	LB137573
CCB07	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	18:46	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	18:46	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	18:46	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	18:46	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	18:46	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	18:46	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	18:46	LB137573
CCB08	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	19:37	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	19:37	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	19:37	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	19:37	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	19:37	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	19:37	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	19:37	LB137573
CCB09	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	20:11	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	20:11	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	20:11	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	20:11	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	20:11	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	20:11	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	20:11	LB137573

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------