

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Scalamandre – Tully JV
Contract: SCAL01
ICS Source: EPA

SDG No.: Q3150
Lab Code: ACE
Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	6.97			-20	20	10/15/2025	14:39	LB137561
	Barium	0.98	6.0	16	-94	106	10/15/2025	14:39	LB137561
	Cadmium	1.63	1.0	163	-5	7	10/15/2025	14:39	LB137561
	Chromium	50.6	52.0	97	42	62	10/15/2025	14:39	LB137561
	Lead	4.81			-12	12	10/15/2025	14:39	LB137561
	Selenium	7.56			-20	20	10/15/2025	14:39	LB137561
	Silver	-6.59			-10	10	10/15/2025	14:39	LB137561
ICSA01	Arsenic	97.5	104	94	88.4	120	10/15/2025	14:52	LB137561
	Barium	445	537	83	437	637	10/15/2025	14:52	LB137561
	Cadmium	914	972	94	826	1120	10/15/2025	14:52	LB137561
	Chromium	504	542	93	460	624	10/15/2025	14:52	LB137561
	Lead	44.5	49.0	91	37	61	10/15/2025	14:52	LB137561
	Selenium	48.7	46.0	106	26	66	10/15/2025	14:52	LB137561
	Silver	184	201	92	170	232	10/15/2025	14:52	LB137561
ICSA	Arsenic	-1.24			-20	20	10/15/2025	14:56	LB137561
	Barium	-39.4	6.0	657	-94	106	10/15/2025	14:56	LB137561
	Cadmium	1.75	1.0	175	-5	7	10/15/2025	14:56	LB137561
	Chromium	48.4	52.0	93	42	62	10/15/2025	14:56	LB137561
	Lead	14.9			-12	12	10/15/2025	14:56	LB137561
	Selenium	-21.2			-20	20	10/15/2025	14:56	LB137561
	Silver	-4.00			-10	10	10/15/2025	14:56	LB137561
ICSA01	Arsenic	137	104	132	88.4	120	10/15/2025	15:00	LB137561
	Barium	509	537	95	437	637	10/15/2025	15:00	LB137561
	Cadmium	975	972	100	826	1120	10/15/2025	15:00	LB137561
	Chromium	571	542	105	460	624	10/15/2025	15:00	LB137561
	Lead	48.6	49.0	99	37	61	10/15/2025	15:00	LB137561
	Selenium	36.4	46.0	79	26	66	10/15/2025	15:00	LB137561
	Silver	183	201	91	170	232	10/15/2025	15:00	LB137561
ICSA01	Arsenic	4.93			-20	20	10/16/2025	12:17	LB137573
	Barium	0.59	6.0	10	-94	106	10/16/2025	12:17	LB137573
	Cadmium	1.20	1.0	120	-5	7	10/16/2025	12:17	LB137573
	Chromium	50.2	52.0	96	42	62	10/16/2025	12:17	LB137573
	Lead	3.50			-12	12	10/16/2025	12:17	LB137573
	Selenium	8.68			-20	20	10/16/2025	12:17	LB137573
	Silver	-7.19			-10	10	10/16/2025	12:17	LB137573
ICSA01	Arsenic	108	104	104	88.4	120	10/16/2025	12:21	LB137573
	Barium	504	537	94	437	637	10/16/2025	12:21	LB137573
	Cadmium	1010	972	104	826	1120	10/16/2025	12:21	LB137573
	Chromium	551	542	102	460	624	10/16/2025	12:21	LB137573
	Lead	49.7	49.0	101	37	61	10/16/2025	12:21	LB137573
	Selenium	55.7	46.0	121	26	66	10/16/2025	12:21	LB137573

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Scalamandre – Tully JV
Contract: SCAL01
ICS Source: EPA

SDG No.: Q3150
Lab Code: ACE
Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Silver	203	201	101	170	232	10/16/2025	12:21	LB137573
ICSA	Arsenic	4.82			-20	20	10/16/2025	12:25	LB137573
	Barium	-96.1	6.0	1602	-94	106	10/16/2025	12:25	LB137573
	Cadmium	-0.22	1.0	22	-5	7	10/16/2025	12:25	LB137573
	Chromium	49.0	52.0	94	42	62	10/16/2025	12:25	LB137573
	Lead	-5.32			-12	12	10/16/2025	12:25	LB137573
	Selenium	-38.2			-20	20	10/16/2025	12:25	LB137573
	Silver	-7.00			-10	10	10/16/2025	12:25	LB137573
ICSAB	Arsenic	114	104	110	88.4	120	10/16/2025	12:29	LB137573
	Barium	550	537	102	437	637	10/16/2025	12:29	LB137573
	Cadmium	998	972	103	826	1120	10/16/2025	12:29	LB137573
	Chromium	606	542	112	460	624	10/16/2025	12:29	LB137573
	Lead	48.4	49.0	99	37	61	10/16/2025	12:29	LB137573
	Selenium	50.5	46.0	110	26	66	10/16/2025	12:29	LB137573
	Silver	199	201	99	170	232	10/16/2025	12:29	LB137573