



Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV41	Mercury	4.04	4.0	101	90 - 110	CV	09/23/2025	11:14	LB137274

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV24	Mercury	5.36	5.0	107	90 - 110	CV	09/23/2025	11:24	LB137274
CCV25	Mercury	4.90	5.0	98	90 - 110	CV	09/23/2025	11:51	LB137274
CCV26	Mercury	4.93	5.0	98	90 - 110	CV	09/23/2025	12:24	LB137274
CCV27	Mercury	4.97	5.0	100	90 - 110	CV	09/23/2025	12:42	LB137274

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3960	4000	99	90 - 110	P	09/22/2025	12:31	LB137271
	Barium	8050	8000	101	90 - 110	P	09/22/2025	12:31	LB137271
	Cadmium	1960	2000	98	90 - 110	P	09/22/2025	12:31	LB137271
	Chromium	803	800	100	90 - 110	P	09/22/2025	12:31	LB137271
	Lead	3850	4000	96	90 - 110	P	09/22/2025	12:31	LB137271
	Selenium	3970	4000	99	90 - 110	P	09/22/2025	12:31	LB137271
	Silver	998	1000	100	90 - 110	P	09/22/2025	12:31	LB137271

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	22.9	20.0	114	80 - 120	P	09/22/2025	12:48	LB137271
	Barium	102	100	102	80 - 120	P	09/22/2025	12:48	LB137271
	Cadmium	6.01	6.0	100	80 - 120	P	09/22/2025	12:48	LB137271
	Chromium	10.4	10.0	104	80 - 120	P	09/22/2025	12:48	LB137271
	Lead	12.3	12.0	102	80 - 120	P	09/22/2025	12:48	LB137271
	Selenium	22.5	20.0	113	80 - 120	P	09/22/2025	12:48	LB137271
	Silver	9.58	10.0	96	80 - 120	P	09/22/2025	12:48	LB137271

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4970	5000	100	90 - 110	P	09/22/2025	13:36	LB137271
	Barium	10100	10000	101	90 - 110	P	09/22/2025	13:36	LB137271
	Cadmium	2460	2500	98	90 - 110	P	09/22/2025	13:36	LB137271
	Chromium	995	1000	100	90 - 110	P	09/22/2025	13:36	LB137271
	Lead	4930	5000	99	90 - 110	P	09/22/2025	13:36	LB137271
	Selenium	4990	5000	100	90 - 110	P	09/22/2025	13:36	LB137271
	Silver	1250	1250	100	90 - 110	P	09/22/2025	13:36	LB137271
CCV02	Arsenic	4890	5000	98	90 - 110	P	09/22/2025	14:48	LB137271
	Barium	10000	10000	100	90 - 110	P	09/22/2025	14:48	LB137271
	Cadmium	2440	2500	98	90 - 110	P	09/22/2025	14:48	LB137271
	Chromium	985	1000	98	90 - 110	P	09/22/2025	14:48	LB137271
	Lead	4880	5000	98	90 - 110	P	09/22/2025	14:48	LB137271
	Selenium	4910	5000	98	90 - 110	P	09/22/2025	14:48	LB137271
	Silver	1240	1250	99	90 - 110	P	09/22/2025	14:48	LB137271
CCV03	Arsenic	4840	5000	97	90 - 110	P	09/22/2025	15:46	LB137271
	Barium	9550	10000	96	90 - 110	P	09/22/2025	15:46	LB137271
	Cadmium	2400	2500	96	90 - 110	P	09/22/2025	15:46	LB137271
	Chromium	965	1000	96	90 - 110	P	09/22/2025	15:46	LB137271
	Lead	4790	5000	96	90 - 110	P	09/22/2025	15:46	LB137271
	Selenium	4880	5000	98	90 - 110	P	09/22/2025	15:46	LB137271
	Silver	1200	1250	96	90 - 110	P	09/22/2025	15:46	LB137271
CCV04	Arsenic	4850	5000	97	90 - 110	P	09/22/2025	16:27	LB137271
	Barium	9630	10000	96	90 - 110	P	09/22/2025	16:27	LB137271
	Cadmium	2400	2500	96	90 - 110	P	09/22/2025	16:27	LB137271
	Chromium	996	1000	100	90 - 110	P	09/22/2025	16:27	LB137271
	Lead	4790	5000	96	90 - 110	P	09/22/2025	16:27	LB137271
	Selenium	4890	5000	98	90 - 110	P	09/22/2025	16:27	LB137271
	Silver	1240	1250	99	90 - 110	P	09/22/2025	16:27	LB137271
CCV05	Arsenic	5060	5000	101	90 - 110	P	09/22/2025	16:43	LB137271
	Barium	9990	10000	100	90 - 110	P	09/22/2025	16:43	LB137271
	Cadmium	2520	2500	101	90 - 110	P	09/22/2025	16:43	LB137271
	Chromium	1010	1000	101	90 - 110	P	09/22/2025	16:43	LB137271
	Lead	5020	5000	100	90 - 110	P	09/22/2025	16:43	LB137271
	Selenium	5090	5000	102	90 - 110	P	09/22/2025	16:43	LB137271

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1260	1250	101	90 - 110	P	09/22/2025	16:43	LB137271
CCV06	Arsenic	4910	5000	98	90 - 110	P	09/22/2025	17:49	LB137271
	Barium	9810	10000	98	90 - 110	P	09/22/2025	17:49	LB137271
	Cadmium	2450	2500	98	90 - 110	P	09/22/2025	17:49	LB137271
	Chromium	976	1000	98	90 - 110	P	09/22/2025	17:49	LB137271
	Lead	4890	5000	98	90 - 110	P	09/22/2025	17:49	LB137271
	Selenium	4960	5000	99	90 - 110	P	09/22/2025	17:49	LB137271
	Silver	1230	1250	98	90 - 110	P	09/22/2025	17:49	LB137271
CCV07	Arsenic	5060	5000	101	90 - 110	P	09/22/2025	18:49	LB137271
	Barium	10200	10000	102	90 - 110	P	09/22/2025	18:49	LB137271
	Cadmium	2540	2500	102	90 - 110	P	09/22/2025	18:49	LB137271
	Chromium	1000	1000	100	90 - 110	P	09/22/2025	18:49	LB137271
	Lead	5070	5000	101	90 - 110	P	09/22/2025	18:49	LB137271
	Selenium	5110	5000	102	90 - 110	P	09/22/2025	18:49	LB137271
	Silver	1250	1250	100	90 - 110	P	09/22/2025	18:49	LB137271

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3850	4000	96	90 - 110	P	09/24/2025	11:34	LB137295
	Barium	7680	8000	96	90 - 110	P	09/24/2025	11:34	LB137295
	Cadmium	1900	2000	95	90 - 110	P	09/24/2025	11:34	LB137295
	Chromium	801	800	100	90 - 110	P	09/24/2025	11:34	LB137295
	Lead	3720	4000	93	90 - 110	P	09/24/2025	11:34	LB137295
	Selenium	3870	4000	97	90 - 110	P	09/24/2025	11:34	LB137295
	Silver	1000	1000	100	90 - 110	P	09/24/2025	11:34	LB137295

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	23.2	20.0	116	80 - 120	P	09/24/2025	11:44	LB137295
	Barium	97.3	100	97	80 - 120	P	09/24/2025	11:44	LB137295
	Cadmium	6.05	6.0	101	80 - 120	P	09/24/2025	11:44	LB137295
	Chromium	10.6	10.0	106	80 - 120	P	09/24/2025	11:44	LB137295
	Lead	11.9	12.0	99	80 - 120	P	09/24/2025	11:44	LB137295
	Selenium	23.9	20.0	120	80 - 120	P	09/24/2025	11:44	LB137295
	Silver	11.8	10.0	118	80 - 120	P	09/24/2025	11:44	LB137295

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4990	5000	100	90 - 110	P	09/24/2025	12:21	LB137295
	Barium	9720	10000	97	90 - 110	P	09/24/2025	12:21	LB137295
	Cadmium	2450	2500	98	90 - 110	P	09/24/2025	12:21	LB137295
	Chromium	1020	1000	102	90 - 110	P	09/24/2025	12:21	LB137295
	Lead	4870	5000	97	90 - 110	P	09/24/2025	12:21	LB137295
	Selenium	5040	5000	101	90 - 110	P	09/24/2025	12:21	LB137295
	Silver	1270	1250	101	90 - 110	P	09/24/2025	12:21	LB137295
CCV02	Arsenic	5130	5000	103	90 - 110	P	09/24/2025	13:40	LB137295
	Barium	9830	10000	98	90 - 110	P	09/24/2025	13:40	LB137295
	Cadmium	2490	2500	100	90 - 110	P	09/24/2025	13:40	LB137295
	Chromium	1040	1000	104	90 - 110	P	09/24/2025	13:40	LB137295
	Lead	4950	5000	99	90 - 110	P	09/24/2025	13:40	LB137295
	Selenium	5200	5000	104	90 - 110	P	09/24/2025	13:40	LB137295
	Silver	1270	1250	102	90 - 110	P	09/24/2025	13:40	LB137295
CCV03	Arsenic	4900	5000	98	90 - 110	P	09/24/2025	14:42	LB137295
	Barium	9580	10000	96	90 - 110	P	09/24/2025	14:42	LB137295
	Cadmium	2400	2500	96	90 - 110	P	09/24/2025	14:42	LB137295
	Chromium	1020	1000	102	90 - 110	P	09/24/2025	14:42	LB137295
	Lead	4760	5000	95	90 - 110	P	09/24/2025	14:42	LB137295
	Selenium	4960	5000	99	90 - 110	P	09/24/2025	14:42	LB137295
	Silver	1230	1250	98	90 - 110	P	09/24/2025	14:42	LB137295
CCV04	Arsenic	5110	5000	102	90 - 110	P	09/24/2025	15:41	LB137295
	Barium	9490	10000	95	90 - 110	P	09/24/2025	15:41	LB137295
	Cadmium	2470	2500	99	90 - 110	P	09/24/2025	15:41	LB137295
	Chromium	1030	1000	103	90 - 110	P	09/24/2025	15:41	LB137295
	Lead	4910	5000	98	90 - 110	P	09/24/2025	15:41	LB137295
	Selenium	5190	5000	104	90 - 110	P	09/24/2025	15:41	LB137295
	Silver	1260	1250	101	90 - 110	P	09/24/2025	15:41	LB137295
CCV05	Arsenic	5200	5000	104	90 - 110	P	09/24/2025	16:31	LB137295
	Barium	9770	10000	98	90 - 110	P	09/24/2025	16:31	LB137295
	Cadmium	2520	2500	101	90 - 110	P	09/24/2025	16:31	LB137295
	Chromium	1050	1000	105	90 - 110	P	09/24/2025	16:31	LB137295
	Lead	5000	5000	100	90 - 110	P	09/24/2025	16:31	LB137295
	Selenium	5310	5000	106	90 - 110	P	09/24/2025	16:31	LB137295

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1280	1250	102	90 - 110	P	09/24/2025	16:31	LB137295
CCV06	Arsenic	5040	5000	101	90 - 110	P	09/24/2025	17:33	LB137295
	Barium	9650	10000	96	90 - 110	P	09/24/2025	17:33	LB137295
	Cadmium	2460	2500	98	90 - 110	P	09/24/2025	17:33	LB137295
	Chromium	1040	1000	104	90 - 110	P	09/24/2025	17:33	LB137295
	Lead	4890	5000	98	90 - 110	P	09/24/2025	17:33	LB137295
	Selenium	5120	5000	102	90 - 110	P	09/24/2025	17:33	LB137295
	Silver	1270	1250	102	90 - 110	P	09/24/2025	17:33	LB137295
CCV07	Arsenic	5030	5000	101	90 - 110	P	09/24/2025	18:37	LB137295
	Barium	9460	10000	95	90 - 110	P	09/24/2025	18:37	LB137295
	Cadmium	2420	2500	97	90 - 110	P	09/24/2025	18:37	LB137295
	Chromium	1020	1000	102	90 - 110	P	09/24/2025	18:37	LB137295
	Lead	4820	5000	96	90 - 110	P	09/24/2025	18:37	LB137295
	Selenium	5120	5000	102	90 - 110	P	09/24/2025	18:37	LB137295
	Silver	1250	1250	100	90 - 110	P	09/24/2025	18:37	LB137295
CCV08	Arsenic	5170	5000	103	90 - 110	P	09/24/2025	19:03	LB137295
	Barium	9540	10000	95	90 - 110	P	09/24/2025	19:03	LB137295
	Cadmium	2480	2500	99	90 - 110	P	09/24/2025	19:03	LB137295
	Chromium	1030	1000	103	90 - 110	P	09/24/2025	19:03	LB137295
	Lead	4920	5000	98	90 - 110	P	09/24/2025	19:03	LB137295
	Selenium	5270	5000	105	90 - 110	P	09/24/2025	19:03	LB137295
	Silver	1270	1250	102	90 - 110	P	09/24/2025	19:03	LB137295

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3970	4000	99	90 - 110	P	09/25/2025	11:20	LB137314
	Barium	7810	8000	98	90 - 110	P	09/25/2025	11:20	LB137314
	Cadmium	1970	2000	98	90 - 110	P	09/25/2025	11:20	LB137314
	Chromium	816	800	102	90 - 110	P	09/25/2025	11:20	LB137314
	Lead	3850	4000	96	90 - 110	P	09/25/2025	11:20	LB137314
	Selenium	4010	4000	100	90 - 110	P	09/25/2025	11:20	LB137314
	Silver	1020	1000	102	90 - 110	P	09/25/2025	11:20	LB137314

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	22.4	20.0	112	80 - 120	P	09/25/2025	11:29	LB137314
	Barium	97.5	100	98	80 - 120	P	09/25/2025	11:29	LB137314
	Cadmium	5.96	6.0	99	80 - 120	P	09/25/2025	11:29	LB137314
	Chromium	10.4	10.0	104	80 - 120	P	09/25/2025	11:29	LB137314
	Lead	10.2	12.0	85	80 - 120	P	09/25/2025	11:29	LB137314
	Selenium	21.6	20.0	108	80 - 120	P	09/25/2025	11:29	LB137314
	Silver	9.87	10.0	99	80 - 120	P	09/25/2025	11:29	LB137314

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4910	5000	98	90 - 110	P	09/25/2025	12:12	LB137314
	Barium	9590	10000	96	90 - 110	P	09/25/2025	12:12	LB137314
	Cadmium	2420	2500	97	90 - 110	P	09/25/2025	12:12	LB137314
	Chromium	981	1000	98	90 - 110	P	09/25/2025	12:12	LB137314
	Lead	4800	5000	96	90 - 110	P	09/25/2025	12:12	LB137314
	Selenium	4970	5000	100	90 - 110	P	09/25/2025	12:12	LB137314
	Silver	1230	1250	98	90 - 110	P	09/25/2025	12:12	LB137314
CCV02	Arsenic	5030	5000	101	90 - 110	P	09/25/2025	13:23	LB137314
	Barium	9690	10000	97	90 - 110	P	09/25/2025	13:23	LB137314
	Cadmium	2460	2500	99	90 - 110	P	09/25/2025	13:23	LB137314
	Chromium	1010	1000	101	90 - 110	P	09/25/2025	13:23	LB137314
	Lead	4880	5000	98	90 - 110	P	09/25/2025	13:23	LB137314
	Selenium	5110	5000	102	90 - 110	P	09/25/2025	13:23	LB137314
	Silver	1260	1250	101	90 - 110	P	09/25/2025	13:23	LB137314
CCV03	Arsenic	4990	5000	100	90 - 110	P	09/25/2025	14:21	LB137314
	Barium	9550	10000	96	90 - 110	P	09/25/2025	14:21	LB137314
	Cadmium	2440	2500	98	90 - 110	P	09/25/2025	14:21	LB137314
	Chromium	981	1000	98	90 - 110	P	09/25/2025	14:21	LB137314
	Lead	4840	5000	97	90 - 110	P	09/25/2025	14:21	LB137314
	Selenium	5080	5000	102	90 - 110	P	09/25/2025	14:21	LB137314
	Silver	1220	1250	98	90 - 110	P	09/25/2025	14:21	LB137314
CCV04	Arsenic	4950	5000	99	90 - 110	P	09/25/2025	15:16	LB137314
	Barium	9580	10000	96	90 - 110	P	09/25/2025	15:16	LB137314
	Cadmium	2440	2500	98	90 - 110	P	09/25/2025	15:16	LB137314
	Chromium	999	1000	100	90 - 110	P	09/25/2025	15:16	LB137314
	Lead	4830	5000	97	90 - 110	P	09/25/2025	15:16	LB137314
	Selenium	5040	5000	101	90 - 110	P	09/25/2025	15:16	LB137314
	Silver	1230	1250	99	90 - 110	P	09/25/2025	15:16	LB137314
CCV05	Arsenic	5170	5000	103	90 - 110	P	09/25/2025	16:16	LB137314
	Barium	9750	10000	98	90 - 110	P	09/25/2025	16:16	LB137314
	Cadmium	2520	2500	101	90 - 110	P	09/25/2025	16:16	LB137314
	Chromium	1040	1000	104	90 - 110	P	09/25/2025	16:16	LB137314
	Lead	4990	5000	100	90 - 110	P	09/25/2025	16:16	LB137314
	Selenium	5280	5000	106	90 - 110	P	09/25/2025	16:16	LB137314

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1290	1250	103	90 - 110	P	09/25/2025	16:16	LB137314
CCV06	Arsenic	5030	5000	100	90 - 110	P	09/25/2025	17:27	LB137314
	Barium	9990	10000	100	90 - 110	P	09/25/2025	17:27	LB137314
	Cadmium	2470	2500	99	90 - 110	P	09/25/2025	17:27	LB137314
	Chromium	1000	1000	100	90 - 110	P	09/25/2025	17:27	LB137314
	Lead	4880	5000	98	90 - 110	P	09/25/2025	17:27	LB137314
	Selenium	5110	5000	102	90 - 110	P	09/25/2025	17:27	LB137314
	Silver	1240	1250	99	90 - 110	P	09/25/2025	17:27	LB137314
CCV07	Arsenic	5120	5000	102	90 - 110	P	09/25/2025	17:40	LB137314
	Barium	9780	10000	98	90 - 110	P	09/25/2025	17:40	LB137314
	Cadmium	2520	2500	101	90 - 110	P	09/25/2025	17:40	LB137314
	Chromium	1040	1000	104	90 - 110	P	09/25/2025	17:40	LB137314
	Lead	4980	5000	100	90 - 110	P	09/25/2025	17:40	LB137314
	Selenium	5210	5000	104	90 - 110	P	09/25/2025	17:40	LB137314
	Silver	1280	1250	102	90 - 110	P	09/25/2025	17:40	LB137314