

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Scalamandre – Tully JV
Contract: SCAL01
ICS Source: EPA

SDG No.: Q3150
Lab Code: ACE
Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	6.86			-20	20	09/22/2025	13:10	LB137271
	Barium	1.22	6.0	20	-94	106	09/22/2025	13:10	LB137271
	Cadmium	-2.33	1.0	233	-5	7	09/22/2025	13:10	LB137271
	Chromium	49.9	52.0	96	42	62	09/22/2025	13:10	LB137271
	Lead	-6.20			-12	12	09/22/2025	13:10	LB137271
	Selenium	-6.14			-20	20	09/22/2025	13:10	LB137271
	Silver	-1.11			-10	10	09/22/2025	13:10	LB137271
ICSAB01	Arsenic	109	104	105	88.4	120	09/22/2025	13:14	LB137271
	Barium	522	537	97	437	637	09/22/2025	13:14	LB137271
	Cadmium	1030	972	106	826	1120	09/22/2025	13:14	LB137271
	Chromium	574	542	106	460	624	09/22/2025	13:14	LB137271
	Lead	45.3	49.0	92	37	61	09/22/2025	13:14	LB137271
	Selenium	49.0	46.0	106	26	66	09/22/2025	13:14	LB137271
	Silver	217	201	108	170	232	09/22/2025	13:14	LB137271
ICSA01	Arsenic	7.04			-20	20	09/24/2025	12:04	LB137295
	Barium	1.38	6.0	23	-94	106	09/24/2025	12:04	LB137295
	Cadmium	1.05	1.0	105	-5	7	09/24/2025	12:04	LB137295
	Chromium	49.2	52.0	95	42	62	09/24/2025	12:04	LB137295
	Lead	0.29			-12	12	09/24/2025	12:04	LB137295
	Selenium	5.58			-20	20	09/24/2025	12:04	LB137295
	Silver	3.73			-10	10	09/24/2025	12:04	LB137295
ICSAB01	Arsenic	109	104	105	88.4	120	09/24/2025	12:08	LB137295
	Barium	477	537	89	437	637	09/24/2025	12:08	LB137295
	Cadmium	959	972	99	826	1120	09/24/2025	12:08	LB137295
	Chromium	549	542	101	460	624	09/24/2025	12:08	LB137295
	Lead	46.7	49.0	95	37	61	09/24/2025	12:08	LB137295
	Selenium	55.6	46.0	121	26	66	09/24/2025	12:08	LB137295
	Silver	214	201	106	170	232	09/24/2025	12:08	LB137295
ICSA01	Arsenic	3.12			-20	20	09/25/2025	11:42	LB137314
	Barium	1.85	6.0	31	-94	106	09/25/2025	11:42	LB137314
	Cadmium	1.36	1.0	136	-5	7	09/25/2025	11:42	LB137314
	Chromium	51.7	52.0	99	42	62	09/25/2025	11:42	LB137314
	Lead	-7.23			-12	12	09/25/2025	11:42	LB137314
	Selenium	10.6			-20	20	09/25/2025	11:42	LB137314
	Silver	-2.79			-10	10	09/25/2025	11:42	LB137314
ICSAB01	Arsenic	112	104	108	88.4	120	09/25/2025	11:46	LB137314
	Barium	497	537	93	437	637	09/25/2025	11:46	LB137314
	Cadmium	1030	972	106	826	1120	09/25/2025	11:46	LB137314
	Chromium	576	542	106	460	624	09/25/2025	11:46	LB137314
	Lead	41.9	49.0	86	37	61	09/25/2025	11:46	LB137314
	Selenium	60.7	46.0	132	26	66	09/25/2025	11:46	LB137314

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Scalamandre – Tully JV
Contract: SCAL01
ICS Source: EPA

SDG No.: Q3150
Lab Code: ACE
Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Silver	220	201	110	170	232	09/25/2025	11:46	LB137314