

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Scalamandre – Tully JV **Level:** LOW **SDG No.:** Q3150
Contract: SCAL01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3157-02 **Client ID:** 20-DEAD-1DUP
Percent Solids for Sample: NA **Duplicate ID** Q3157-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	698		723		4		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	12.5	J	26.4	J	71		P
Lead	ug/L	20	20.3	J	15.6	J	26		P
Mercury	ug/L	20	2.00	U	0.77	J	200.0		CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Scalamandre – Tully JV **Level:** LOW **SDG No.:** Q3150
Contract: SCAL01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3157-02MS **Client ID:** 20-DEAD-1MSD
Percent Solids for Sample: NA **Duplicate ID** Q3157-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3990		4160		4		P
Barium	ug/L	20	1540		1690		9		P
Cadmium	ug/L	20	957		1010		5		P
Chromium	ug/L	20	2050		2140		4		P
Lead	ug/L	20	4600		4850		5		P
Mercury	ug/L	20	49.6		45.8		8		CV
Selenium	ug/L	20	10100		10500		4		P
Silver	ug/L	20	371		385		4		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”