

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169773BS							
Arsenic	ug/L	4000	3900		98	80 - 120	P
Barium	ug/L	1000	953		95	80 - 120	P
Cadmium	ug/L	1000	941		94	80 - 120	P
Chromium	ug/L	2000	1970		98	80 - 120	P
Lead	ug/L	5000	4730		95	80 - 120	P
Selenium	ug/L	10000	9970		100	80 - 120	P
Silver	ug/L	380	373		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q3150

Contract: SCAL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169781BS Mercury	ug/L	4.0	4.67		117	80 - 120	CV