

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCAL01
Lab Code: ACE **SDG NO.:** Q3150
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113919.D **Time Analyzed:** 18:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.643	4.843	491.900	500.000	-1.6
Aroclor-1016-2	4.760	4.661	4.861	498.360	500.000	-0.3
Aroclor-1016-3	4.817	4.718	4.918	478.980	500.000	-4.2
Aroclor-1016-4	4.936	4.837	5.037	486.290	500.000	-2.7
Aroclor-1016-5	5.192	5.093	5.293	567.270	500.000	13.5
Aroclor-1260-1	6.227	6.129	6.329	510.180	500.000	2.0
Aroclor-1260-2	6.416	6.319	6.519	480.240	500.000	-4.0
Aroclor-1260-3	6.781	6.683	6.883	477.220	500.000	-4.6
Aroclor-1260-4	7.039	6.941	7.141	535.900	500.000	7.2
Aroclor-1260-5	7.283	7.185	7.385	541.560	500.000	8.3
Decachlorobiphenyl	8.668	8.570	8.770	43.100	50.000	-13.8
Tetrachloro-m-xylene	3.659	3.560	3.760	52.680	50.000	5.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCAL01
Lab Code: ACE **SDG NO.:** Q3150
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113919.D **Time Analyzed:** 18:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.623	4.823	488.360	500.000	-2.3
Aroclor-1016-2	4.740	4.641	4.841	483.290	500.000	-3.3
Aroclor-1016-3	4.914	4.816	5.016	492.170	500.000	-1.6
Aroclor-1016-4	4.957	4.858	5.058	488.570	500.000	-2.3
Aroclor-1016-5	5.168	5.070	5.270	510.810	500.000	2.2
Aroclor-1260-1	6.193	6.096	6.296	463.780	500.000	-7.2
Aroclor-1260-2	6.382	6.284	6.484	460.500	500.000	-7.9
Aroclor-1260-3	6.533	6.435	6.635	433.520	500.000	-13.3
Aroclor-1260-4	7.001	6.903	7.103	445.890	500.000	-10.8
Aroclor-1260-5	7.244	7.146	7.346	442.880	500.000	-11.4
Decachlorobiphenyl	8.612	8.514	8.714	48.500	50.000	-3.0
Tetrachloro-m-xylene	3.650	3.551	3.751	51.130	50.000	2.3