

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** SCAL01  
**Lab Code:** ACE **SDG NO.:** Q3150  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 09/22/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113934.D **Time Analyzed:** 23:42

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.741 | 4.643     | 4.843 | 496.170            | 500.000           | -0.8  |
| Aroclor-1016-2       | 4.759 | 4.661     | 4.861 | 507.120            | 500.000           | 1.4   |
| Aroclor-1016-3       | 4.816 | 4.718     | 4.918 | 486.730            | 500.000           | -2.7  |
| Aroclor-1016-4       | 4.936 | 4.837     | 5.037 | 498.110            | 500.000           | -0.4  |
| Aroclor-1016-5       | 5.191 | 5.093     | 5.293 | 542.650            | 500.000           | 8.5   |
| Aroclor-1260-1       | 6.227 | 6.129     | 6.329 | 513.560            | 500.000           | 2.7   |
| Aroclor-1260-2       | 6.417 | 6.319     | 6.519 | 481.860            | 500.000           | -3.6  |
| Aroclor-1260-3       | 6.781 | 6.683     | 6.883 | 504.340            | 500.000           | 0.9   |
| Aroclor-1260-4       | 7.040 | 6.941     | 7.141 | 567.770            | 500.000           | 13.6  |
| Aroclor-1260-5       | 7.284 | 7.185     | 7.385 | 554.800            | 500.000           | 11.0  |
| Decachlorobiphenyl   | 8.668 | 8.570     | 8.770 | 42.410             | 50.000            | -15.2 |
| Tetrachloro-m-xylene | 3.659 | 3.560     | 3.760 | 51.890             | 50.000            | 3.8   |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** SCAL01  
**Lab Code:** ACE **SDG NO.:** Q3150  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 09/22/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113934.D **Time Analyzed:** 23:42

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.721 | 4.623     | 4.823 | 485.040            | 500.000           | -3.0  |
| Aroclor-1016-2       | 4.738 | 4.641     | 4.841 | 487.440            | 500.000           | -2.5  |
| Aroclor-1016-3       | 4.913 | 4.816     | 5.016 | 507.830            | 500.000           | 1.6   |
| Aroclor-1016-4       | 4.957 | 4.858     | 5.058 | 495.660            | 500.000           | -0.9  |
| Aroclor-1016-5       | 5.167 | 5.070     | 5.270 | 529.080            | 500.000           | 5.8   |
| Aroclor-1260-1       | 6.193 | 6.096     | 6.296 | 484.270            | 500.000           | -3.1  |
| Aroclor-1260-2       | 6.382 | 6.284     | 6.484 | 474.670            | 500.000           | -5.1  |
| Aroclor-1260-3       | 6.532 | 6.435     | 6.635 | 430.720            | 500.000           | -13.9 |
| Aroclor-1260-4       | 7.001 | 6.903     | 7.103 | 446.490            | 500.000           | -10.7 |
| Aroclor-1260-5       | 7.244 | 7.146     | 7.346 | 441.530            | 500.000           | -11.7 |
| Decachlorobiphenyl   | 8.612 | 8.514     | 8.714 | 48.110             | 50.000            | -3.8  |
| Tetrachloro-m-xylene | 3.650 | 3.551     | 3.751 | 50.260             | 50.000            | 0.5   |