

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCAL01
Lab Code: ACE **SDG NO.:** Q3150
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113948.D **Time Analyzed:** 05:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.742	4.643	4.843	507.230	500.000	1.4
Aroclor-1016-2	4.760	4.661	4.861	525.810	500.000	5.2
Aroclor-1016-3	4.817	4.718	4.918	520.560	500.000	4.1
Aroclor-1016-4	4.936	4.837	5.037	512.640	500.000	2.5
Aroclor-1016-5	5.192	5.093	5.293	540.270	500.000	8.1
Aroclor-1260-1	6.226	6.129	6.329	556.490	500.000	11.3
Aroclor-1260-2	6.417	6.319	6.519	532.460	500.000	6.5
Aroclor-1260-3	6.782	6.683	6.883	530.960	500.000	6.2
Aroclor-1260-4	7.039	6.941	7.141	593.180	500.000	18.6
Aroclor-1260-5	7.283	7.185	7.385	593.790	500.000	18.8
Decachlorobiphenyl	8.668	8.570	8.770	48.670	50.000	-2.7
Tetrachloro-m-xylene	3.659	3.560	3.760	53.560	50.000	7.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCAL01
Lab Code: ACE **SDG NO.:** Q3150
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113948.D **Time Analyzed:** 05:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.623	4.823	496.850	500.000	-0.6
Aroclor-1016-2	4.738	4.641	4.841	497.320	500.000	-0.5
Aroclor-1016-3	4.913	4.816	5.016	515.340	500.000	3.1
Aroclor-1016-4	4.956	4.858	5.058	504.390	500.000	0.9
Aroclor-1016-5	5.167	5.070	5.270	534.610	500.000	6.9
Aroclor-1260-1	6.193	6.096	6.296	492.560	500.000	-1.5
Aroclor-1260-2	6.382	6.284	6.484	485.240	500.000	-3.0
Aroclor-1260-3	6.532	6.435	6.635	460.360	500.000	-7.9
Aroclor-1260-4	7.001	6.903	7.103	476.490	500.000	-4.7
Aroclor-1260-5	7.243	7.146	7.346	472.800	500.000	-5.4
Decachlorobiphenyl	8.611	8.514	8.714	51.530	50.000	3.1
Tetrachloro-m-xylene	3.650	3.551	3.751	51.760	50.000	3.5