

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCAL01
Lab Code: ACE **SDG NO.:** Q3150
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/10/2025 09/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070933.D **Time Analyzed:** 13:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.481	4.381	4.581	513.770	500.000	2.8
Aroclor-1016-2	4.499	4.400	4.600	501.390	500.000	0.3
Aroclor-1016-3	4.557	4.457	4.657	505.140	500.000	1.0
Aroclor-1016-4	4.647	4.548	4.748	504.520	500.000	0.9
Aroclor-1016-5	4.931	4.832	5.032	496.020	500.000	-0.8
Aroclor-1260-1	6.021	5.921	6.121	475.250	500.000	-5.0
Aroclor-1260-2	6.276	6.177	6.377	463.260	500.000	-7.3
Aroclor-1260-3	6.627	6.528	6.728	478.450	500.000	-4.3
Aroclor-1260-4	6.841	6.742	6.942	465.380	500.000	-6.9
Aroclor-1260-5	7.147	7.047	7.247	466.690	500.000	-6.7
Decachlorobiphenyl	8.497	8.399	8.599	47.760	50.000	-4.5
Tetrachloro-m-xylene	3.401	3.302	3.502	51.930	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCAL01
Lab Code: ACE **SDG NO.:** Q3150
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/10/2025 09/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070933.D **Time Analyzed:** 13:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.778	3.678	3.878	468.310	500.000	-6.3
Aroclor-1016-2	3.794	3.694	3.894	461.740	500.000	-7.7
Aroclor-1016-3	3.954	3.854	4.054	455.180	500.000	-9.0
Aroclor-1016-4	4.003	3.903	4.103	454.740	500.000	-9.1
Aroclor-1016-5	4.197	4.097	4.297	467.020	500.000	-6.6
Aroclor-1260-1	5.190	5.090	5.290	437.300	500.000	-12.5
Aroclor-1260-2	5.379	5.280	5.480	441.930	500.000	-11.6
Aroclor-1260-3	5.519	5.420	5.620	436.480	500.000	-12.7
Aroclor-1260-4	5.979	5.880	6.080	419.480	500.000	-16.1
Aroclor-1260-5	6.224	6.126	6.326	418.970	500.000	-16.2
Decachlorobiphenyl	7.533	7.436	7.636	38.730	50.000	-22.5
Tetrachloro-m-xylene	2.778	2.678	2.878	47.130	50.000	-5.7