

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCAL01
Lab Code: ACE **SDG NO.:** Q3150
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/10/2025 09/10/2025

Client Sample No.: CCAL06 **Date Analyzed:** 09/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070944.D **Time Analyzed:** 18:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.481	4.381	4.581	556.890	500.000	11.4
Aroclor-1016-2	4.499	4.400	4.600	526.870	500.000	5.4
Aroclor-1016-3	4.557	4.457	4.657	536.580	500.000	7.3
Aroclor-1016-4	4.647	4.548	4.748	538.050	500.000	7.6
Aroclor-1016-5	4.932	4.832	5.032	534.980	500.000	7.0
Aroclor-1260-1	6.021	5.921	6.121	458.190	500.000	-8.4
Aroclor-1260-2	6.277	6.177	6.377	508.510	500.000	1.7
Aroclor-1260-3	6.628	6.528	6.728	526.050	500.000	5.2
Aroclor-1260-4	6.841	6.742	6.942	481.440	500.000	-3.7
Aroclor-1260-5	7.146	7.047	7.247	453.090	500.000	-9.4
Decachlorobiphenyl	8.498	8.399	8.599	43.890	50.000	-12.2
Tetrachloro-m-xylene	3.401	3.302	3.502	54.310	50.000	8.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCAL01
Lab Code: ACE **SDG NO.:** Q3150
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/10/2025 09/10/2025

Client Sample No.: CCAL06 **Date Analyzed:** 09/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070944.D **Time Analyzed:** 18:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.777	3.678	3.878	478.160	500.000	-4.4
Aroclor-1016-2	3.793	3.694	3.894	471.340	500.000	-5.7
Aroclor-1016-3	3.953	3.854	4.054	465.970	500.000	-6.8
Aroclor-1016-4	4.002	3.903	4.103	465.760	500.000	-6.8
Aroclor-1016-5	4.197	4.097	4.297	476.280	500.000	-4.7
Aroclor-1260-1	5.189	5.090	5.290	469.880	500.000	-6.0
Aroclor-1260-2	5.378	5.280	5.480	459.590	500.000	-8.1
Aroclor-1260-3	5.520	5.420	5.620	448.810	500.000	-10.2
Aroclor-1260-4	5.980	5.880	6.080	414.370	500.000	-17.1
Aroclor-1260-5	6.224	6.126	6.326	409.830	500.000	-18.0
Decachlorobiphenyl	7.534	7.436	7.636	34.360	50.000	-31.3
Tetrachloro-m-xylene	2.778	2.678	2.878	49.680	50.000	-0.6