

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092225\
 Data File : P0113934.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 23:42
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/23/2025
 Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:46:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 16 02:49:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.659	3.650	457.7E6	268.4E6	51.895	50.258
2)	SA Decachlor...	8.668	8.612	283.7E6	100.4E6	42.407	48.113
Target Compounds							
3)	L1 AR-1016-1	4.741	4.721	150.6E6	88395257	496.166	485.045
4)	L1 AR-1016-2	4.759	4.738	230.8E6	134.9E6	507.117	487.444
5)	L1 AR-1016-3	4.816	4.913	145.6E6	75840889	486.727	507.828
6)	L1 AR-1016-4	4.936	4.957	119.7E6	59831065	498.108	495.662
7)	L1 AR-1016-5	5.191	5.167	130.3E6	79397156	542.646m	529.081
31)	L7 AR-1260-1	6.227	6.193	251.7E6	121.7E6	513.564	484.267
32)	L7 AR-1260-2	6.417	6.382	369.0E6	156.9E6	481.863	474.675
33)	L7 AR-1260-3	6.781	6.532	339.3E6	120.8E6	504.337	430.715
34)	L7 AR-1260-4	7.040	7.001	245.7E6	85373420	567.769m	446.487
35)	L7 AR-1260-5	7.284	7.244	628.5E6	188.0E6	554.798	441.531

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092225\
Data File : PO113934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 23:42
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/23/2025
Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:46:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 16 02:49:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

