

Surrogate Summary

SDG No.: **Q3150**

Client: **Scalamandre – Tully JV**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113719.D	PIBLK-PO113719.D	Tetrachloro-m-xyl	1	20	17.0	85		60	140
		Decachlorobiphen	1	20	17.3	86		60	140
		Tetrachloro-m-xyl	2	20	17.0	85		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO113923.D	PIBLK-PO113923.D	Tetrachloro-m-xyl	1	20	18.2	91		60	140
		Decachlorobiphen	1	20	16.1	81		60	140
		Tetrachloro-m-xyl	2	20	17.7	88		60	140
		Decachlorobiphen	2	20	18.3	92		60	140
Q3150-01	MER Rd Con Ed	Tetrachloro-m-xyl	1	20	10.3	51		32	144
		Decachlorobiphen	1	20	13.5	67		32	175
		Tetrachloro-m-xyl	2	20	12.3	61		32	144
		Decachlorobiphen	2	20	13.8	69		32	175
Q3150-03	Mid Site Grid 5-7	Tetrachloro-m-xyl	1	20	12.1	60		32	144
		Decachlorobiphen	1	20	4.77	24	*	32	175
		Tetrachloro-m-xyl	2	20	11.3	56		32	144
		Decachlorobiphen	2	20	7.44	37		32	175
I.BLK-PO113938.D	PIBLK-PO113938.D	Tetrachloro-m-xyl	1	20	18.1	90		60	140
		Decachlorobiphen	1	20	16.4	82		60	140
		Tetrachloro-m-xyl	2	20	17.4	87		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO113952.D	PIBLK-PO113952.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	18.2	91		60	140
		Tetrachloro-m-xyl	2	20	17.5	87		60	140
		Decachlorobiphen	2	20	18.9	95		60	140
PB169771BL	PB169771BL	Tetrachloro-m-xyl	1	20	19.7	99		32	144
		Decachlorobiphen	1	20	20.2	101		32	175
		Tetrachloro-m-xyl	2	20	19.1	95		32	144
		Decachlorobiphen	2	20	20.8	104		32	175
PB169771BS	PB169771BS	Tetrachloro-m-xyl	1	20	19.4	97		32	144
		Decachlorobiphen	1	20	20.6	103		32	175
		Tetrachloro-m-xyl	2	20	18.5	93		32	144
		Decachlorobiphen	2	20	21.6	108		32	175
I.BLK-PO113958.D	PIBLK-PO113958.D	Tetrachloro-m-xyl	1	20	18.2	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	17.5	87		60	140
		Decachlorobiphen	2	20	18.9	95		60	140
I.BLK-PQ070788.D	PIBLK-PQ070788.D	Tetrachloro-m-xyl	1	20	17.2	86		60	140
		Decachlorobiphen	1	20	18.9	94		60	140
		Tetrachloro-m-xyl	2	20	17.6	88		60	140
		Decachlorobiphen	2	20	18.5	93		60	140
I.BLK-PQ070934.D	PIBLK-PQ070934.D	Tetrachloro-m-xyl	1	20	16.9	85		60	140

Surrogate Summary

SDG No.: Q3150

Client: Scalamandre – Tully JV

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PQ070934.D	PIBLK-PQ070934.D	Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	15.8	79		60	140
		Decachlorobiphen	2	20	14.9	74		60	140
Q3152-01MS	EO-02-091925MS	Tetrachloro-m-xyl	1	20	14.4	72		32	144
		Decachlorobiphen	1	20	13.3	67		32	175
		Tetrachloro-m-xyl	2	20	12.7	64		32	144
		Decachlorobiphen	2	20	9.14	46		32	175
Q3152-01MSD	EO-02-091925MSD	Tetrachloro-m-xyl	1	20	15.0	75		32	144
		Decachlorobiphen	1	20	12.9	64		32	175
		Tetrachloro-m-xyl	2	20	13.1	66		32	144
		Decachlorobiphen	2	20	8.88	44		32	175
I.BLK-PQ070945.D	PIBLK-PQ070945.D	Tetrachloro-m-xyl	1	20	17.3	86		60	140
		Decachlorobiphen	1	20	16.0	80		60	140
		Tetrachloro-m-xyl	2	20	15.4	77		60	140
		Decachlorobiphen	2	20	12.7	64		60	140