

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3150</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Scalamandre – Tully JV</u>	Datafile :	<u>PO113954.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169771BS (Column 1)	AR1016	166.5	156	ug/kg	94				71	120	
	AR1260	166.5	163	ug/kg	98				65	130	
PB169771BS (Column 2)	AR1016	166.5	146	ug/kg	88				71	120	
	AR1260	166.5	139	ug/kg	83				65	130	