

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BP092525

SequenceID : BP092525

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. Compound n-Nitrosodimethylamine was marginally low & Benzaldehyde was biased high side in SSTDCCC040 (BP025855.D) and it is not present in the samples under the sequence.	_____	_____✓_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____	_____✓_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

Nitrobenzene-d5 surrogate was marginally biased low in PB169801BL,BS & PB169791TB & hence, the data will be used for the hard copies. Surrogates recoveries failed in the sample Q3168-06 due to presence of non target hydrocarbons in the sample. The samples Q3150-02 & Q3167-01 were analyzed with their respective dilution due to viscous matrix difficult to inject otherwise.

Jagrut

Analyst

09/25/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BP092525

SequenceID : BP092525 NA NO YES