

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3150

Client ID: SSTDCCC040

Date Analyzed: 09/24/2025

Lab File ID: BP025840.D

Time Analyzed: 10:58

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	828956	17.177	1048550	21.63	1127310	24.989
UPPER LIMIT	1657910	17.677	2097100	22.13	2254620	25.489
LOWER LIMIT	414478	16.677	524275	21.13	563655	24.489
EPA SAMPLE NO.						
Mid Site Grid 5-7	945408	17.18	975743	21.63	1079830	24.99

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.