

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: SCAL01
 Lab Code: ACE SDG No.: Q3150
 Instrument ID: MSVOA_W Calibration Date/Time: 09/19/2025 09:38
 Lab File ID: VW032204.D Init. Calib. Date(s): 08/26/2025 08/26/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:39 12:34
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	48.12	3.751999	20