



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092225\
Data File : VW032222.D
Acq On : 22 Sep 2025 08:04
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Title : SW846 8260
Last Update : Wed Aug 27 04:13:40 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.3	18760	PASS
75	95	30	60	56.0	49387	PASS
95	95	100	100	100.0	88120	PASS
96	95	5	9	6.5	5757	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	68.1	60048	PASS
175	174	5	9	8.1	4872	PASS
176	174	95	101	98.9	59408	PASS
177	176	5	9	6.7	3952	PASS