



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092525\
Data File : VW032257.D
Acq On : 25 Sep 2025 10:01
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Title : SW846 8260
Last Update : Fri Sep 26 07:11:42 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1750

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	28296	PASS
75	95	30	60	51.0	75976	PASS
95	95	100	100	100.0	149099	PASS
96	95	5	9	6.5	9671	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	66.7	99464	PASS
175	174	5	9	7.4	7318	PASS
176	174	95	101	96.6	96075	PASS
177	176	5	9	6.2	5983	PASS