

Surrogate Summary

SDG No.: **Q3150**

Client: **Scalamandre – Tully JV**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3150-02	MER Rd Con Ed	1,2-Dichloroethane-d4	50	49.8	100		63	155
		Dibromofluoromethane	50	46.2	92		70	134
		Toluene-d8	50	40.0	80		74	123
		4-Bromofluorobenzene	50	41.4	83		17	146
Q3150-02RE	MER Rd Con EdRE	1,2-Dichloroethane-d4	50	63.5	127		63	155
		Dibromofluoromethane	50	54.8	110		70	134
		Toluene-d8	50	41.1	82		74	123
		4-Bromofluorobenzene	50	29.9	60		17	146
Q3150-04	Mid Site Grid 5-7	1,2-Dichloroethane-d4	50	52.5	105		63	155
		Dibromofluoromethane	50	46.3	93		70	134
		Toluene-d8	50	43.1	86		74	123
		4-Bromofluorobenzene	50	37.2	74		17	146
Q3150-04RE	Mid Site Grid 5-7RE	1,2-Dichloroethane-d4	50	56.4	113		63	155
		Dibromofluoromethane	50	52.5	105		70	134
		Toluene-d8	50	48.0	96		74	123
		4-Bromofluorobenzene	50	37.5	75		17	146
VW0919SBL01	VW0919SBL01	1,2-Dichloroethane-d4	50	58.7	117		63	155
		Dibromofluoromethane	50	48.5	97		70	134
		Toluene-d8	50	44.7	89		74	123
		4-Bromofluorobenzene	50	47.4	95		17	146
VW0919SBS01	VW0919SBS01	1,2-Dichloroethane-d4	50	52.9	106		63	155
		Dibromofluoromethane	50	52.9	106		70	134
		Toluene-d8	50	52.6	105		74	123
		4-Bromofluorobenzene	50	50.3	101		17	146
VW0919SBSD0	VW0919SBSD01	1,2-Dichloroethane-d4	50	53.3	107		63	155
		Dibromofluoromethane	50	51.3	103		70	134
		Toluene-d8	50	51.5	103		74	123
		4-Bromofluorobenzene	50	49.5	99		17	146
VW0922SBL01	VW0922SBL01	1,2-Dichloroethane-d4	50	65.7	131		63	155
		Dibromofluoromethane	50	55.7	111		70	134
		Toluene-d8	50	56.2	112		74	123
		4-Bromofluorobenzene	50	51.3	103		17	146
VW0922SBS01	VW0922SBS01	1,2-Dichloroethane-d4	50	49.6	99		63	155
		Dibromofluoromethane	50	51.6	103		70	134
		Toluene-d8	50	50.6	101		74	123
		4-Bromofluorobenzene	50	49.4	99		17	146
VW0930SBL01	VW0930SBL01	1,2-Dichloroethane-d4	50	57.0	114		63	155
		Dibromofluoromethane	50	50.2	100		70	134
		Toluene-d8	50	51.1	102		74	123
		4-Bromofluorobenzene	50	46.1	92		17	146
VW0930SBS01	VW0930SBS01	1,2-Dichloroethane-d4	50	49.9	100		63	155
		Dibromofluoromethane	50	46.7	93		70	134
		Toluene-d8	50	46.6	93		74	123
		4-Bromofluorobenzene	50	46.6	93		17	146
VW1002SBL01	VW1002SBL01	1,2-Dichloroethane-d4	50	59.9	120		63	155
		Dibromofluoromethane	50	52.7	105		70	134
		Toluene-d8	50	53.3	107		74	123
		4-Bromofluorobenzene	50	50.8	102		17	146
VW1002SBS01	VW1002SBS01	1,2-Dichloroethane-d4	50	52.2	104		63	155
		Dibromofluoromethane	50	50.1	100		70	134

Surrogate Summary

SDG No.: Q3150

Client: Scalamandre – Tully JV

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
VW1002SBS01	VW1002SBS01	Toluene-d8	50	51.8	104		74	123
		4-Bromofluorobenzene	50	51.5	103		17	146