

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3150</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Scalamandre – Tully JV</u>	Datafile :	<u>VW032207.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0919SBSD01	Dichlorodifluoromethane	20	25.6	ug/Kg	128	12		64	136	20
	Chloromethane	20	27.0	ug/Kg	135	8		52	151	20
	Vinyl chloride	20	27.4	ug/Kg	137	4		56	148	20
	Bromomethane	20	26.6	ug/Kg	133	5		58	141	20
	Chloroethane	20	28.3	ug/Kg	142	10	*	69	130	20
	Trichlorofluoromethane	20	25.4	ug/Kg	127	24	*	69	134	20
	1,1,2-Trichlorotrifluoroethane	20	24.5	ug/Kg	123	2		81	123	20
	1,1-Dichloroethene	20	25.3	ug/Kg	127	3	*	79	121	20
	Acetone	100	150	ug/Kg	150	6		40	171	20
	Carbon disulfide	20	22.0	ug/Kg	110	7		59	130	20
	Methyl tert-butyl Ether	20	22.6	ug/Kg	113	9		77	129	20
	Methyl Acetate	20	20.7	ug/Kg	104	9		69	149	20
	Methylene Chloride	20	24.0	ug/Kg	120	13		72	131	20
	trans-1,2-Dichloroethene	20	22.8	ug/Kg	114	9		80	123	20
	1,1-Dichloroethane	20	23.3	ug/Kg	117	10		82	123	20
	Cyclohexane	20	22.1	ug/Kg	111	5		76	122	20
	2-Butanone	100	120	ug/Kg	120	9		69	131	20
	Carbon Tetrachloride	20	21.9	ug/Kg	110	2		76	129	20
	cis-1,2-Dichloroethene	20	22.2	ug/Kg	111	9		82	123	20
	Bromochloromethane	20	23.3	ug/Kg	117	3		80	127	20
	Chloroform	20	23.5	ug/Kg	117	9		82	125	20
	1,1,1-Trichloroethane	20	23.5	ug/Kg	117	10		80	126	20
	Methylcyclohexane	20	21.5	ug/Kg	108	5		77	123	20
	Benzene	20	22.6	ug/Kg	113	5		84	121	20
	1,2-Dichloroethane	20	23.2	ug/Kg	116	7		81	126	20
	Trichloroethene	20	22.3	ug/Kg	112	8		83	122	20
	1,2-Dichloropropane	20	22.9	ug/Kg	115	6		83	122	20
	Bromodichloromethane	20	22.5	ug/Kg	113	6		82	123	20
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	10		70	135	20
	Toluene	20	21.9	ug/Kg	110	4		83	122	20
	t-1,3-Dichloropropene	20	21.7	ug/Kg	109	5		78	124	20
	cis-1,3-Dichloropropene	20	22.1	ug/Kg	111	7		81	122	20
	1,1,2-Trichloroethane	20	23.0	ug/Kg	115	8		82	125	20
	2-Hexanone	100	120	ug/Kg	120	9		66	138	20
	Dibromochloromethane	20	22.3	ug/Kg	112	13		79	125	20
	1,2-Dibromoethane	20	21.6	ug/Kg	108	5		80	125	20
	Tetrachloroethene	20	21.1	ug/Kg	106	5		83	125	20
	Chlorobenzene	20	21.7	ug/Kg	109	5		84	122	20
	Ethyl Benzene	20	22.1	ug/Kg	111	7		82	124	20
	m/p-Xylenes	40	45.3	ug/Kg	113	6		83	124	20
	o-Xylene	20	21.7	ug/Kg	109	7		83	123	20
	Styrene	20	21.8	ug/Kg	109	5		82	124	20
	Bromoform	20	20.9	ug/Kg	104	4		75	127	20
	Isopropylbenzene	20	20.8	ug/Kg	104	3		82	124	20
	1,1,2,2-Tetrachloroethane	20	22.4	ug/Kg	112	6		77	127	20
	1,3-Dichlorobenzene	20	21.5	ug/Kg	108	5		83	122	20
	1,4-Dichlorobenzene	20	22.4	ug/Kg	112	11		84	121	20
	1,2-Dichlorobenzene	20	22.2	ug/Kg	111	8		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3150</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Scalamandre – Tully JV</u>	Datafile :	<u>VW032207.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0919SBSD01	1,2-Dibromo-3-Chloropropane	20	22.0	ug/Kg	110	11		66	134	20
	1,2,4-Trichlorobenzene	20	20.7	ug/Kg	104	8		78	127	20
	1,2,3-Trichlorobenzene	20	20.3	ug/Kg	102	4		70	137	20