

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCAL01
 Lab Code: ACE SDG NO.: Q3150
 Lab File ID: VW032204.D Date Analyzed: 09/19/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	193605	7.95	337702	8.85	309889	11.63
UPPER LIMIT	387210	8.453	675404	9.349	619778	12.129
LOWER LIMIT	96802.5	7.453	168851	8.349	154945	11.129
EPA SAMPLE NO.						
MER Rd Con Ed	64103 *	7.96	128658 *	8.85	115897 *	11.64
VW0919SBL01	128568	7.96	281974	8.85	287563	11.63
VW0919SBS01	187746	7.96	322734	8.85	298059	11.63
VW0919SBSD01	180121	7.96	317195	8.86	296209	11.64

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.