

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCAL01
 Lab Code: ACE SDG NO.: Q3150
 Lab File ID: VW032223.D Date Analyzed: 09/22/2025
 Instrument ID: MSVOA_W Time Analyzed: 08:47
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	160129	13.55				
UPPER LIMIT	320258	14.05				
LOWER LIMIT	80064.5	13.05				
EPA SAMPLE NO.						
MER Rd Con EdRE	4044 *	13.56				
VW0922SBL01	151656	13.56				
VW0922SBS01	149741	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.