

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: SCAL01

Lab Code: ACE

SDG NO.: Q3150

Lab File ID: VW032222.D

BFB Injection Date: 09/22/2025

Instrument ID: MSVOA_W

BFB Injection Time: 08:04

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.3
75	30.0 - 60.0% of mass 95	56
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 100.0% of mass 95	68.1
175	5.0 - 9.0% of mass 174	5.5 (8.1) 1
176	95.0 - 101.0% of mass 174	67.4 (98.9) 1
177	5.0 - 9.0% of mass 176	4.5 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VW032223.D	09/22/2025	08:47
VW0922SBL01	VW0922SBL01	VW032224.D	09/22/2025	10:27
VW0922SBS01	VW0922SBS01	VW032225.D	09/22/2025	11:13
MER Rd Con EdRE	Q3150-02RE	VW032227.D	09/22/2025	12:11