

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL092225

Review By	Mahesh Dadoda	Review On	9/24/2025 2:18:14 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/24/2025 2:26:00 AM		
SubDirectory	VL092225	HP Acquire Method	TO15AIR	HP Processing Method	VL092225AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2676
Initial Calibration Stds	AP2673,AP2674,AP2675
CCC	AP2673
Internal Standard/PEM	AP2676
ICV/I.BLK	AP2671
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042949.D	22 Sep 2025 08:55		SY/MD	Ok
2	VSTDICCC010	VSTDICCC010	VL042950.D	22 Sep 2025 09:39		SY/MD	Ok,M
3	VSTDICC002	VSTDICC002	VL042951.D	22 Sep 2025 10:14	LR-20	SY/MD	Ok,M
4	VSTDICC001	VSTDICC001	VL042952.D	22 Sep 2025 10:48		SY/MD	Ok,M
5	VSTDICC0.5	VSTDICC0.5	VL042953.D	22 Sep 2025 11:22		SY/MD	Ok,M
6	VSTDICC0.1	VSTDICC0.1	VL042954.D	22 Sep 2025 11:56		SY/MD	Ok,M
7	VSTDICC0.03	VSTDICC0.03	VL042955.D	22 Sep 2025 12:30		SY/MD	Ok,M
8	VSTDICC015	VSTDICC015	VL042956.D	22 Sep 2025 13:05		SY/MD	Ok,M
9	VSTDICV010	ICVVL092225	VL042957.D	22 Sep 2025 14:15	ICV failed for com.#13	SY/MD	Ok,M
10	VL0922ABL01	VL0922ABL01	VL042958.D	22 Sep 2025 15:06		SY/MD	Ok
11	VL0922ABS01	VL0922ABS01	VL042959.D	22 Sep 2025 15:52		SY/MD	Ok,M
12	Q3111-01	SV1	VL042960.D	22 Sep 2025 16:53	Need 40X	SY/MD	Dilution
13	Q3111-01DL	SV1DL	VL042961.D	22 Sep 2025 17:28		SY/MD	Ok,M
14	Q3112-01	SV1	VL042962.D	22 Sep 2025 18:04		SY/MD	Ok,M
15	Q3112-01DL	SV1DL	VL042963.D	22 Sep 2025 18:39	Not Required	SY/MD	Not Ok
16	Q3112-02	SV2	VL042964.D	22 Sep 2025 19:14	Need 40X	SY/MD	Dilution
17	Q3112-02DL	SV2DL	VL042965.D	22 Sep 2025 19:49		SY/MD	Ok
18	Q3112-03	SV3	VL042966.D	22 Sep 2025 20:24		SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL092225

Review By	Mahesh Dadoda	Review On	9/24/2025 2:18:14 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/24/2025 2:26:00 AM		
SubDirectory	VL092225	HP Acquire Method	TO15AIR	HP Processing Method	VL092225AIR.M
STD. NAME	STD REF.#				
Tune/Reschk	AP2676				
Initial Calibration Stds	AP2673,AP2674,AP2675				
CCC	AP2673				
Internal Standard/PEM	AP2676				
ICV/I.BLK	AP2671				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3112-03DL	SV3DL	VL042967.D	22 Sep 2025 20:59	Not Required	SY/MD	Not Ok
20	VIBLK	VIBLK	VL042968.D	22 Sep 2025 21:35		SY/MD	Ok
21	Q3130-02	131-1A-2	VL042969.D	22 Sep 2025 22:11		SY/MD	Ok,M
22	Q3130-02DUP	131-1A-2DUP	VL042970.D	22 Sep 2025 22:48		SY/MD	Ok,M
23	Q3130-02DL	131-1A-2DL	VL042971.D	22 Sep 2025 23:23	Not Required	SY/MD	Not Ok
24	Q3130-01	131-1A-1	VL042972.D	23 Sep 2025 00:00		SY/MD	Ok,M
25	Q3130-01DL	131-1A-1DL	VL042973.D	23 Sep 2025 00:35	Not Required	SY/MD	Not Ok
26	Q3131-01	141-1A-1	VL042974.D	23 Sep 2025 01:11	Need 10X	SY/MD	Dilution
27	Q3131-01DL	141-1A-1DL	VL042975.D	23 Sep 2025 01:46		SY/MD	Ok,M
28	Q3131-02	141-1A-2	VL042976.D	23 Sep 2025 02:22	Need 10X	SY/MD	Dilution
29	Q3131-02DL	141-1A-2DL	VL042977.D	23 Sep 2025 02:57		SY/MD	Ok,M
30	Q3132-01	151-1A-1	VL042978.D	23 Sep 2025 03:33	Need 10X	SY/MD	Dilution
31	Q3132-01DL	151-1A-1DL	VL042979.D	23 Sep 2025 04:08		SY/MD	Ok,M
32	Q3132-02	151-1A-2	VL042980.D	23 Sep 2025 04:44		SY/MD	Ok,M
33	Q3132-02DL	151-1A-2DL	VL042981.D	23 Sep 2025 05:19	Not Required	SY/MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL092425

Review By	Semsettin Yesilyurt	Review On	9/25/2025 11:06:05 AM		
Supervise By	Mahesh Dadoda	Supervise On	9/26/2025 1:44:26 AM		
SubDirectory	VL092425	HP Acquire Method	TO15AIR	HP Processing Method	VL092225AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2676
Initial Calibration Stds	AP2673,AP2674,AP2675
CCC	AP2673
Internal Standard/PEM	AP2676
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL043000.D	24 Sep 2025 08:20		SY/MD	Ok
2	VSTDCCC010	VSTDCCC010	VL043001.D	24 Sep 2025 09:03		SY/MD	Ok,M
3	VL0924ABL01	VL0924ABL01	VL043002.D	24 Sep 2025 09:51		SY/MD	Ok
4	VL0924ABS01	VL0924ABS01	VL043003.D	24 Sep 2025 10:47		SY/MD	Ok,M
5	Q3154-02DL	SV2DL	VL043004.D	24 Sep 2025 11:29	Not Required	SY/MD	Not Ok
6	Q3154-01DL	SV1DL	VL043005.D	24 Sep 2025 12:03	Not Required	SY/MD	Not Ok
7	Q3154-02DL	SV2DL	VL043006.D	24 Sep 2025 12:44	Not Required	SY/MD	Not Ok
8	Q3154-02	SV2	VL043007.D	24 Sep 2025 13:31		SY/MD	Ok,M
9	Q3154-02DUP	SV2DUP	VL043008.D	24 Sep 2025 14:06		SY/MD	Ok,M
10	Q3154-01	SV1	VL043009.D	24 Sep 2025 14:40	Run @ Lower Dilution	SY/MD	Not Ok
11	Q3154-01	SV1	VL043010.D	24 Sep 2025 15:39		SY/MD	Ok,M
12	VIBLK	VIBLK	VL043011.D	24 Sep 2025 16:14		SY/MD	Ok
13	Q2893-15 0.45	MDL-AIR-03-QT3-2025	VL043012.D	24 Sep 2025 16:49		SY/MD	Ok,M
14	Q2893-15 0.40	MDL-AIR-03-QT3-2025	VL043013.D	24 Sep 2025 17:24		SY/MD	Ok,M
15	Q2893-15 0.1	MDL-AIR-03-QT3-2025	VL043014.D	24 Sep 2025 17:59		SY/MD	Ok,M
16	Q2893-15 0.03	MDL-AIR-03-QT3-2025	VL043015.D	24 Sep 2025 19:49		SY/MD	Ok,M

M : Manual Integration