

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q3155

Project: Tanner G. Duckrey Public School

RunNo.: LB137254

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10	10	100	90-110	09/09/2025
Chloride	mg/L	3	3	100	90-110	09/09/2025
Fluoride	mg/L	2	2	100	90-110	09/09/2025
Nitrite	mg/L	3	3	100	90-110	09/09/2025
Nitrate	mg/L	2.5	2.5	100	90-110	09/09/2025
Sulfate	mg/L	14.8	15	99	90-110	09/09/2025
Orthophosphate as P	mg/L	5	5	100	90-110	09/09/2025
Sample ID: CCV1						
Bromide	mg/L	9.9	10	99	90-110	09/22/2025
Chloride	mg/L	2.9	3	97	90-110	09/22/2025
Fluoride	mg/L	2	2	100	90-110	09/22/2025
Nitrite	mg/L	2.9	3	97	90-110	09/22/2025
Nitrate	mg/L	2.4	2.5	96	90-110	09/22/2025
Sulfate	mg/L	14.6	15	97	90-110	09/22/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	09/22/2025
Sample ID: CCV2						
Bromide	mg/L	9.9	10	99	90-110	09/22/2025
Chloride	mg/L	2.9	3	97	90-110	09/22/2025
Fluoride	mg/L	2	2	100	90-110	09/22/2025
Nitrite	mg/L	2.9	3	97	90-110	09/22/2025
Nitrate	mg/L	2.4	2.5	96	90-110	09/22/2025
Sulfate	mg/L	14.6	15	97	90-110	09/22/2025
Orthophosphate as P	mg/L	5	5	100	90-110	09/22/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q3155

Project: Tanner G. Duckrey Public School

RunNo.: LB137261

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	09/22/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	09/22/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	09/22/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	09/22/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q3155

Project: Tanner G. Duckrey Public School

RunNo.: LB137276

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.496	0.5	99	90-110	09/23/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	09/23/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.492	0.5	98	90-110	09/23/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	09/23/2025