

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q3155

Project: Tanner G. Duckrey Public School

Run No.: LB137254

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137254BSS							
Bromide	mg/Kg	200	196		98	1	90-110	09/22/2025
Chloride	mg/Kg	60	57.8		96	1	90-110	09/22/2025
Fluoride	mg/Kg	40	39.2		98	1	90-110	09/22/2025
Nitrite	mg/Kg	60	58.4		97	1	90-110	09/22/2025
Nitrate	mg/Kg	50	47.6		95	1	90-110	09/22/2025
Sulfate	mg/Kg	300	290		97	1	90-110	09/22/2025
Orthophosphate as P	mg/Kg	100	99.2		99	1	90-110	09/22/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q3155

Project: Tanner G. Duckrey Public School

Run No.: LB137261

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169778BS							
Ammonia as N	mg/Kg	50	48.6		97	1	90-110	09/22/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q3155

Project: Tanner G. Duckrey Public School

Run No.: LB137276

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169786BS							
Hexavalent Chromium	mg/Kg	20	19.8		99	1	84-110	09/23/2025