

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** POWE02

**Lab Code:** ACE

**SDG NO.:** Q3155

**Continuing Calib Date:** 09/23/2025

**Initial Calibration Date(s):** 09/15/2025 09/15/2025

**Continuing Calib Time:** 00:56

**Initial Calibration Time(s):** 09:34 18:27

**GC Column:** ZB-MR1 **ID:** 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56                   | 3.76 | 0.00       |
| Decachlorobiphenyl   | 8.67       | 8.67      | 8.57                   | 8.77 | 0.00       |

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** POWE02

**Lab Code:** ACE

**SDG NO.:** Q3155

**Continuing Calib Date:** 09/23/2025

**Initial Calibration Date(s):** 09/15/2025 09/15/2025

**Continuing Calib Time:** 00:56

**Initial Calibration Time(s):** 09:34 18:27

**GC Column:** ZB-MR2 **ID:** 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Tetrachloro-m-xylene | 3.65       | 3.65      | 3.55      | 3.75 | 0.00       |
| Decachlorobiphenyl   | 8.61       | 8.61      | 8.51      | 8.71 | 0.00       |