

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3155

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/15/2025 09/15/2025

Continuing Calib Time: 06:29

Initial Calibration Time(s): 09:34 18:27

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.67	8.67	8.57	8.77	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3155

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/15/2025 09/15/2025

Continuing Calib Time: 06:29

Initial Calibration Time(s): 09:34 18:27

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.61	8.61	8.51	8.71	0.00