

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3155
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113907.D **Time Analyzed:** 13:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.643	4.843	511.880	500.000	2.4
Aroclor-1016-2	4.759	4.661	4.861	513.760	500.000	2.8
Aroclor-1016-3	4.816	4.718	4.918	498.030	500.000	-0.4
Aroclor-1016-4	4.935	4.837	5.037	502.570	500.000	0.5
Aroclor-1016-5	5.191	5.093	5.293	550.220	500.000	10.0
Aroclor-1260-1	6.227	6.129	6.329	514.910	500.000	3.0
Aroclor-1260-2	6.416	6.319	6.519	497.920	500.000	-0.4
Aroclor-1260-3	6.781	6.683	6.883	509.790	500.000	2.0
Aroclor-1260-4	7.040	6.941	7.141	555.930	500.000	11.2
Aroclor-1260-5	7.284	7.185	7.385	561.650	500.000	12.3
Decachlorobiphenyl	8.667	8.570	8.770	46.230	50.000	-7.5
Tetrachloro-m-xylene	3.659	3.560	3.760	53.850	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3155
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113907.D **Time Analyzed:** 13:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.623	4.823	488.230	500.000	-2.4
Aroclor-1016-2	4.738	4.641	4.841	492.580	500.000	-1.5
Aroclor-1016-3	4.913	4.816	5.016	496.450	500.000	-0.7
Aroclor-1016-4	4.956	4.858	5.058	493.470	500.000	-1.3
Aroclor-1016-5	5.167	5.070	5.270	509.880	500.000	2.0
Aroclor-1260-1	6.193	6.096	6.296	471.840	500.000	-5.6
Aroclor-1260-2	6.382	6.284	6.484	459.060	500.000	-8.2
Aroclor-1260-3	6.532	6.435	6.635	442.950	500.000	-11.4
Aroclor-1260-4	7.001	6.903	7.103	457.730	500.000	-8.5
Aroclor-1260-5	7.244	7.146	7.346	455.040	500.000	-9.0
Decachlorobiphenyl	8.612	8.514	8.714	49.130	50.000	-1.7
Tetrachloro-m-xylene	3.650	3.551	3.751	51.590	50.000	3.2